

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041323\
 Data File : P0094210.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 13 Apr 2023 18:46
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 13 23:09:09 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.419	3.623	170.1E6	65211748	45.494	51.210
2) SA Decachlor...	10.245	8.643	115.1E6	49312878	51.462	49.323
Target Compounds						
16) L4 AR-1242-1	5.595	4.706	42115071	16247344	482.505	527.981
17) L4 AR-1242-2	5.618	4.724	62033301	23855248	466.537	518.659
18) L4 AR-1242-3	5.680	4.900	39751166	12090195	468.280	498.889
19) L4 AR-1242-4	5.779	4.985	29767791	13705236	476.429	488.188
20) L4 AR-1242-5	6.523	5.508	30041998	15444503	446.189	541.705



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041323\
Data File : PO094210.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 13 Apr 2023 18:46
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 13 23:09:09 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.419 min
Delta R.T.: -0.003 min
Response: 170068204
Conc: 45.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.623 min
Delta R.T.: -0.004 min
Response: 65211748
Conc: 51.21 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.245 min
Delta R.T.: -0.006 min
Response: 115117222
Conc: 51.46 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.643 min
Delta R.T.: -0.008 min
Response: 49312878
Conc: 49.32 ng/ml



#16 AR-1242-1


R.T.: 5.595 min
Delta R.T.: -0.003 min
Response: 42115071
Conc: 482.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.706 min
Delta R.T.: -0.004 min
Response: 16247344
Conc: 527.98 ng/ml




#17 AR-1242-2

R.T.: 5.618 min
Delta R.T.: -0.002 min
Response: 62033301
Conc: 466.54 ng/ml



#17 AR-1242-2


R.T.: 4.724 min
Delta R.T.: -0.004 min
Response: 23855248
Conc: 518.66 ng/ml



#18 AR-1242-3


R.T.: 5.680 min
Delta R.T.: -0.002 min
Response: 39751166
Conc: 468.28 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.900 min
Delta R.T.: -0.004 min
Response: 12090195
Conc: 498.89 ng/ml




#19 AR-1242-4

R.T.: 5.779 min
Delta R.T.: -0.003 min
Response: 29767791
Conc: 476.43 ng/ml



#19 AR-1242-4


R.T.: 4.985 min
Delta R.T.: -0.004 min
Response: 13705236
Conc: 488.19 ng/ml



#20 AR-1242-5

R.T.: 6.523 min
Delta R.T.: -0.002 min
Response: 30041998
Conc: 446.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.508 min
Delta R.T.: -0.004 min
Response: 15444503
Conc: 541.71 ng/ml