

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041425\
 Data File : P0110422.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 14 Apr 2025 19:23
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 14 23:55:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.690	3.687	495.2E6	277.4E6	56.605	55.591
2) SA Decachlor...	8.738	8.689	276.4E6	79954139	35.008	41.535
Target Compounds						
26) L6 AR-1254-1	5.589	5.568	317.2E6	164.9E6	568.517	546.837
27) L6 AR-1254-2	5.738	5.715	278.4E6	141.8E6	573.201	535.656
28) L6 AR-1254-3	6.144	6.118	412.8E6	216.0E6	527.090	535.405
29) L6 AR-1254-4	6.373	6.346	224.6E6	124.2E6	460.713	540.687
30) L6 AR-1254-5	6.793	6.763	268.9E6	156.8E6	383.656	471.824



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041425\
Data File : PO110422.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 14 Apr 2025 19:23
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 14 23:55:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.690 min
Delta R.T.: 0.002 min
Response: 495244816
Conc: 56.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 277388931
Conc: 55.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.738 min
Delta R.T.: 0.006 min
Response: 276385032
Conc: 35.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.689 min
Delta R.T.: 0.005 min
Response: 79954139
Conc: 41.53 ng/ml



#26 AR-1254-1


R.T.: 5.589 min
Delta R.T.: 0.002 min
Response: 317226540
Conc: 568.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.568 min
Delta R.T.: 0.002 min
Response: 164870932
Conc: 546.84 ng/ml




#27 AR-1254-2

R.T.: 5.738 min
Delta R.T.: 0.002 min
Response: 278405087
Conc: 573.20 ng/ml



#27 AR-1254-2


R.T.: 5.715 min
Delta R.T.: 0.002 min
Response: 141780246
Conc: 535.66 ng/ml



#28 AR-1254-3


R.T.: 6.144 min
Delta R.T.: 0.002 min
Response: 412828624
Conc: 527.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.118 min
Delta R.T.: 0.001 min
Response: 215990957
Conc: 535.40 ng/ml




#29 AR-1254-4

R.T.: 6.373 min
Delta R.T.: 0.002 min
Response: 224646309
Conc: 460.71 ng/ml



#29 AR-1254-4


R.T.: 6.346 min
Delta R.T.: 0.002 min
Response: 124202881
Conc: 540.69 ng/ml



#30 AR-1254-5

R.T.: 6.793 min
Delta R.T.: 0.003 min
Response: 268904696
Conc: 383.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.763 min
Delta R.T.: 0.002 min
Response: 156783320
Conc: 471.82 ng/ml