

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041723\
 Data File : P0094285.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2023 18:30
 Operator : YP/AJ
 Sample : 02366-07
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 20230335-COMP-13

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 17 22:01:21 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.418	3.620	68714227	26213708	18.382	20.585
2) SA Decachlor...	10.261	8.641	43490371	20063683	19.442	20.068

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041723\
Data File : PO094285.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2023 18:30
Operator : YP/AJ
Sample : 02366-07
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
20230335-COMP-13

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 17 22:01:21 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.418 min
Delta R.T.: -0.004 min
Response: 68714227
Conc: 18.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
20230335-COMP-13




#1 Tetrachloro-m-xylene

R.T.: 3.620 min
Delta R.T.: -0.006 min
Response: 26213708
Conc: 20.59 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: 0.010 min
Response: 43490371
Conc: 19.44 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.641 min
Delta R.T.: -0.010 min
Response: 20063683
Conc: 20.07 ng/ml