

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041723\
 Data File : P0094300.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 17 Apr 2023 23:43
 Operator : YP/AJ
 Sample : PB152147BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 PB152147BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 05:26:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.417	3.621	66111024	24822824	17.685	19.493
2) SA Decachlor...	10.261	8.640	45659083	21008989	20.412	21.013

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041723\
Data File : PO094300.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 17 Apr 2023 23:43
Operator : YP/AJ
Sample : PB152147BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
PB152147BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 05:26:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.417 min
Delta R.T.: -0.005 min
Response: 66111024
Conc: 17.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
PB152147BL




#1 Tetrachloro-m-xylene

R.T.: 3.621 min
Delta R.T.: -0.006 min
Response: 24822824
Conc: 19.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.261 min
Delta R.T.: 0.010 min
Response: 45659083
Conc: 20.41 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.640 min
Delta R.T.: -0.011 min
Response: 21008989
Conc: 21.01 ng/ml