

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041823\
 Data File : P0094360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2023 21:42
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 19 06:39:01 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0040323.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 04 06:20:55 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.415	3.618	144.3E6	57849400	38.609	45.428
2) SA Decachlor...	10.257	8.637	92333676	38201914	41.277	38.210
Target Compounds						
26) L6 AR-1254-1	6.458	5.502	48094970	30119733	354.612	440.646
27) L6 AR-1254-2	6.679	5.650	75397939	26183790	385.720	429.910
28) L6 AR-1254-3	7.048	6.054	76882575	38852306	416.160	426.536
29) L6 AR-1254-4	7.338	6.284	51252021	21006438	505.383	493.164
30) L6 AR-1254-5	7.762	6.704	57075312	31433167	420.896	435.644



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041823\
Data File : PO094360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2023 21:42
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 19 06:39:01 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO040323.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 04 06:20:55 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: -0.007 min
Response: 144327308
Conc: 38.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: -0.009 min
Response: 57849400
Conc: 45.43 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.257 min
Delta R.T.: 0.006 min
Response: 92333676
Conc: 41.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.637 min
Delta R.T.: -0.014 min
Response: 38201914
Conc: 38.21 ng/ml



#26 AR-1254-1


R.T.: 6.458 min
Delta R.T.: -0.002 min
Response: 48094970
Conc: 354.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.502 min
Delta R.T.: -0.010 min
Response: 30119733
Conc: 440.65 ng/ml




#27 AR-1254-2

R.T.: 6.679 min
Delta R.T.: -0.003 min
Response: 75397939
Conc: 385.72 ng/ml



#27 AR-1254-2


R.T.: 5.650 min
Delta R.T.: -0.010 min
Response: 26183790
Conc: 429.91 ng/ml



#28 AR-1254-3


R.T.: 7.048 min
Delta R.T.: -0.002 min
Response: 76882575
Conc: 416.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.054 min
Delta R.T.: -0.011 min
Response: 38852306
Conc: 426.54 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: 0.000 min
Response: 51252021
Conc: 505.38 ng/ml



#29 AR-1254-4


R.T.: 6.284 min
Delta R.T.: -0.010 min
Response: 21006438
Conc: 493.16 ng/ml



#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: 0.000 min
Response: 57075312
Conc: 420.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.704 min
Delta R.T.: -0.010 min
Response: 31433167
Conc: 435.64 ng/ml