

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041825\
 Data File : P0110526.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 18 Apr 2025 11:42
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 18 13:35:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	496.9E6	271.2E6	56.797	54.345
2) SA Decachlor...	8.732	8.684	396.8E6	96767011	50.263	50.269
Target Compounds						
26) L6 AR-1254-1	5.586	5.565	293.1E6	152.8E6	525.293	506.910
27) L6 AR-1254-2	5.735	5.712	256.5E6	133.7E6	528.128	505.019
28) L6 AR-1254-3	6.140	6.115	414.2E6	207.5E6	528.856	514.433
29) L6 AR-1254-4	6.369	6.342	257.3E6	120.1E6	527.586	522.938
30) L6 AR-1254-5	6.789	6.760	360.7E6	169.0E6	514.636	508.492



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041825\
Data File : PO110526.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 18 Apr 2025 11:42
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 18 13:35:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: -0.001 min
Response: 496918333
Conc: 56.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.002 min
Response: 271171622
Conc: 54.35 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 396822955
Conc: 50.26 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 96767011
Conc: 50.27 ng/ml



#26 AR-1254-1


R.T.: 5.586 min
Delta R.T.: -0.001 min
Response: 293107878
Conc: 525.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.565 min
Delta R.T.: -0.001 min
Response: 152832844
Conc: 506.91 ng/ml




#27 AR-1254-2

R.T.: 5.735 min
Delta R.T.: -0.002 min
Response: 256512892
Conc: 528.13 ng/ml



#27 AR-1254-2


R.T.: 5.712 min
Delta R.T.: -0.001 min
Response: 133670993
Conc: 505.02 ng/ml



#28 AR-1254-3


R.T.: 6.140 min
Delta R.T.: -0.002 min
Response: 414211363
Conc: 528.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: -0.002 min
Response: 207530559
Conc: 514.43 ng/ml




#29 AR-1254-4

R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 257253719
Conc: 527.59 ng/ml



#29 AR-1254-4


R.T.: 6.342 min
Delta R.T.: -0.002 min
Response: 120125615
Conc: 522.94 ng/ml



#30 AR-1254-5

R.T.: 6.789 min
Delta R.T.: -0.001 min
Response: 360708207
Conc: 514.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: -0.001 min
Response: 168967704
Conc: 508.49 ng/ml