

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0041923.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 20 04:54:55 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094397.D 1000 =P0094395.D 750 =P0094396.D
 250 =P0094398.D 50 =P0094399.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	3.461	3.347	3.430	3.478	3.303	3.404	E6	2.23
2)	SA	Decachlorobiphenyl	3.975	3.697	3.851	4.131	3.889	3.909	E6	4.09
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00
6)	L1	AR-1016-4						0.000		-1.00
7)	L1	AR-1016-5						0.000		-1.00
8)	L2	AR-1221-1						0.000		-1.00
9)	L2	AR-1221-2						0.000		-1.00
10)	L2	AR-1221-3						0.000		-1.00
11)	L3	AR-1232-1						0.000		-1.00
12)	L3	AR-1232-2						0.000		-1.00
13)	L3	AR-1232-3						0.000		-1.00
14)	L3	AR-1232-4						0.000		-1.00
15)	L3	AR-1232-5						0.000		-1.00
16)	L4	AR-1242-1						0.000		-1.00
17)	L4	AR-1242-2						0.000		-1.00
18)	L4	AR-1242-3						0.000		-1.00
19)	L4	AR-1242-4						0.000		-1.00
20)	L4	AR-1242-5						0.000		-1.00
21)	L5	AR-1248-1						0.000		-1.00
22)	L5	AR-1248-2						0.000		-1.00
23)	L5	AR-1248-3						0.000		-1.00
24)	L5	AR-1248-4						0.000		-1.00
25)	L5	AR-1248-5						0.000		-1.00
26)	L6	AR-1254-1						0.000		-1.00
27)	L6	AR-1254-2						0.000		-1.00
28)	L6	AR-1254-3						0.000		-1.00
29)	L6	AR-1254-4						0.000		-1.00
30)	L6	AR-1254-5						0.000		-1.00
31)	L7	AR-1260-1						0.000		-1.00
32)	L7	AR-1260-2						0.000		-1.00
33)	L7	AR-1260-3						0.000		-1.00
34)	L7	AR-1260-4						0.000		-1.00
35)	L7	AR-1260-5						0.000		-1.00
36)	L8	AR-1262-1						0.000		-1.00
37)	L8	AR-1262-2						0.000		-1.00
38)	L8	AR-1262-3						0.000		-1.00
39)	L8	AR-1262-4						0.000		-1.00
40)	L8	AR-1262-5						0.000		-1.00
41)	L9	AR-1268-1	3.192	3.047	3.115	3.284	3.220	3.171	E5	2.91
42)	L9	AR-1268-2	2.853	2.718	2.785	2.944	2.806	2.821	E5	2.98
43)	L9	AR-1268-3	2.666	2.541	2.616	2.753	2.603	2.636	E5	3.00
44)	L9	AR-1268-4	8.849	8.354	8.599	8.862	7.515	8.436	E4	6.58
45)	L9	AR-1268-5	8.067	7.796	7.966	8.172	7.999	8.000	E5	1.73

Signal #2 Calibration Files

500 =P0094397.D 1000 =P0094395.D 750 =P0094396.D
 250 =P0094398.D 50 =P0094399.D

Compound			500	1000	750	250	50	Avg	%RSD	
1)	SA	Tetrachloro-m-xylene	1.353	1.291	1.330	1.371	1.281	1.325	E6	2.94
2)	SA	Decachlorobiphenyl	1.817	1.672	1.745	1.919	2.008	1.832	E6	7.33
3)	L1	AR-1016-1						0.000		-1.00
4)	L1	AR-1016-2						0.000		-1.00
5)	L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0041923.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 20 04:54:55 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094397.D 1000 =P0094395.D 750 =P0094396.D
 250 =P0094398.D 50 =P0094399.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.661	1.560	1.615	1.725	1.751	1.662 E5	4.71
42) L9	AR-1268-2	1.550	1.456	1.511	1.617	1.564	1.540 E5	3.91
43) L9	AR-1268-3	1.373	1.279	1.329	1.438	1.443	1.372 E5	5.12
44) L9	AR-1268-4	4.628	4.280	4.449	4.852	4.452	4.532 E4	4.79
45) L9	AR-1268-5	3.736	3.570	3.666	3.837	3.842	3.730 E5	3.10

(#) = Out of Range