

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041923\
 Data File : P0094403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 02:08
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041923AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 04:36:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 04:35:58 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.415	3.617	175.5E6	68726576	51.820	51.930
2) SA Decachlor...	10.251	8.632	116.0E6	54428012	51.878	51.014
Target Compounds						
26) L6 AR-1254-1	6.456	5.500	62683268	35331952	506.915	505.508
27) L6 AR-1254-2	6.678	5.648	91534430	31547613	505.594	503.833
28) L6 AR-1254-3	7.047	6.052	87058095	46636664	510.323	504.455
29) L6 AR-1254-4	7.335	6.281	48263084	21839568	512.128	515.435
30) L6 AR-1254-5	7.760	6.700	61883457	38477557	518.338	513.379



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041923\
Data File : PO094403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 02:08
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041923AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 04:36:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 04:35:58 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: 0.000 min
Response: 175484191
Conc: 51.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041923AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: 0.000 min
Response: 68726576
Conc: 51.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.251 min
Delta R.T.: 0.000 min
Response: 115984334
Conc: 51.88 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.632 min
Delta R.T.: -0.001 min
Response: 54428012
Conc: 51.01 ng/ml



#26 AR-1254-1


R.T.: 6.456 min
Delta R.T.: 0.000 min
Response: 62683268
Conc: 506.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041923AR1254




#26 AR-1254-1

R.T.: 5.500 min
Delta R.T.: 0.000 min
Response: 35331952
Conc: 505.51 ng/ml




#27 AR-1254-2

R.T.: 6.678 min
Delta R.T.: 0.000 min
Response: 91534430
Conc: 505.59 ng/ml



#27 AR-1254-2


R.T.: 5.648 min
Delta R.T.: 0.000 min
Response: 31547613
Conc: 503.83 ng/ml



#28 AR-1254-3


R.T.: 7.047 min
Delta R.T.: 0.000 min
Response: 87058095
Conc: 510.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041923AR1254




#28 AR-1254-3

R.T.: 6.052 min
Delta R.T.: 0.000 min
Response: 46636664
Conc: 504.45 ng/ml




#29 AR-1254-4

R.T.: 7.335 min
Delta R.T.: 0.000 min
Response: 48263084
Conc: 512.13 ng/ml



#29 AR-1254-4


R.T.: 6.281 min
Delta R.T.: 0.000 min
Response: 21839568
Conc: 515.43 ng/ml



#30 AR-1254-5

R.T.: 7.760 min
Delta R.T.: 0.000 min
Response: 61883457
Conc: 518.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041923AR1254



#30 AR-1254-5

R.T.: 6.700 min
Delta R.T.: 0.000 min
Response: 38477557
Conc: 513.38 ng/ml