

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042023\
 Data File : P0094410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 09:54
 Operator : YP/AJ
 Sample : 02390-11 2X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SPC-3WC-23-04-17

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/21/2023
 Supervised By :Ankita Jodhani 04/21/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 20 21:29:38 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 05:15:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.617	26844522	9324378	8.344m	7.219m
2) SA Decachlor...	10.250	8.632	15398513	6685714	7.204	6.713
Target Compounds						
16) L4 AR-1242-1	5.593	4.699	20421250	7812761	258.325	245.430
17) L4 AR-1242-2	5.617	4.717	49803564	15997951	410.354	334.927
18) L4 AR-1242-3	5.686	4.895	15452824	8551123	197.181	337.873 #
19) L4 AR-1242-4	5.774	4.977	19037371	19638316	326.496m	675.917 #
20) L4 AR-1242-5	6.522	5.501	51974978	39360931	869.285	1338.688 #
26) L6 AR-1254-1	6.456	5.501	65958612	39360931	533.402	563.152
27) L6 AR-1254-2	6.677	5.649	160.6E6	21284626	887.087	339.927 #
28) L6 AR-1254-3	7.058	6.060	378.9E6	30914207	2220.836	334.390m#
29) L6 AR-1254-4	7.335	6.287	31313418	35897766	332.272	847.222 #
30) L6 AR-1254-5	7.759	6.701	67250703	35339829	563.294m	471.514

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042023\
Data File : PO094410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 09:54
Operator : YP/AJ
Sample : 02390-11 2X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPC-3WC-23-04-17

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/21/2023
Supervised By :Ankita Jodhani 04/21/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 20 21:29:38 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 05:15:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

