

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042023\
 Data File : P0094451.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Apr 2023 23:02
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:10:32 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 05:15:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.617	177.2E6	68084700	55.076	52.709
2) SA Decachlor...	10.250	8.631	121.6E6	55130676	56.890	55.352
Target Compounds						
26) L6 AR-1254-1	6.455	5.500	60317498	37532204	487.783	536.988
27) L6 AR-1254-2	6.677	5.647	92142339	33371962	508.952	532.969
28) L6 AR-1254-3	7.047	6.051	92107013	50030647	539.919	541.167
29) L6 AR-1254-4	7.336	6.281	58484816	26359028	620.593	622.098
30) L6 AR-1254-5	7.759	6.700	71095595	42161255	595.499	562.528



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042023\
Data File : PO094451.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Apr 2023 23:02
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:10:32 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 05:15:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.004 min
Response: 177181067
Conc: 55.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: -0.005 min
Response: 68084700
Conc: 52.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.250 min
Delta R.T.: -0.009 min
Response: 121599166
Conc: 56.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.631 min
Delta R.T.: -0.008 min
Response: 55130676
Conc: 55.35 ng/ml



#26 AR-1254-1


R.T.: 6.455 min
Delta R.T.: -0.001 min
Response: 60317498
Conc: 487.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.500 min
Delta R.T.: -0.001 min
Response: 37532204
Conc: 536.99 ng/ml




#27 AR-1254-2

R.T.: 6.677 min
Delta R.T.: 0.000 min
Response: 92142339
Conc: 508.95 ng/ml



#27 AR-1254-2


R.T.: 5.647 min
Delta R.T.: -0.002 min
Response: 33371962
Conc: 532.97 ng/ml



#28 AR-1254-3


R.T.: 7.047 min
Delta R.T.: -0.002 min
Response: 92107013
Conc: 539.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.051 min
Delta R.T.: -0.002 min
Response: 50030647
Conc: 541.17 ng/ml




#29 AR-1254-4

R.T.: 7.336 min
Delta R.T.: 0.000 min
Response: 58484816
Conc: 620.59 ng/ml



#29 AR-1254-4


R.T.: 6.281 min
Delta R.T.: -0.002 min
Response: 26359028
Conc: 622.10 ng/ml



#30 AR-1254-5

R.T.: 7.759 min
Delta R.T.: -0.001 min
Response: 71095595
Conc: 595.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.700 min
Delta R.T.: -0.002 min
Response: 42161255
Conc: 562.53 ng/ml