

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042023\
 Data File : P0094461.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 01:53
 Operator : YP/AJ
 Sample : 02424-09 10X
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 KMA8975-2-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 05:15:24 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 05:15:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.413	3.616	5746120	2222758	1.786	1.721
2) SA Decachlor...	10.251	8.631	4359037	2151888	2.039	2.161

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042023\
Data File : PO094461.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 01:53
Operator : YP/AJ
Sample : 02424-09 10X
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
KMA8975-2-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 05:15:24 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 05:15:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.413 min
Delta R.T.: -0.005 min
Response: 5746120
Conc: 1.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
KMA8975-2-1




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.005 min
Response: 2222758
Conc: 1.72 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.251 min
Delta R.T.: -0.007 min
Response: 4359037
Conc: 2.04 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.631 min
Delta R.T.: -0.008 min
Response: 2151888
Conc: 2.16 ng/ml