

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042123\
 Data File : P0094490.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Apr 2023 12:38
 Operator : YP/AJ
 Sample : 02423-11DL 4X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 SPE-1WC-23-04-18DL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 21 21:45:45 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 20 05:15:06 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.417	3.619	19269698	7040176	5.990	5.450
2) SA Decachlor...	10.258	8.637	14606261	6098284	6.833	6.123
Target Compounds						
26) L6 AR-1254-1	6.460	5.504	51650693	31986949	417.695	457.650
27) L6 AR-1254-2	6.681	5.652	94430057	25353057	521.588	404.902
28) L6 AR-1254-3	7.050	6.056	90708305	47929505	531.720	518.439
29) L6 AR-1254-4	7.339	6.286	37412472	18905888	396.990	446.197
30) L6 AR-1254-5	7.764	6.708	90400322	158.8E6	757.196	2119.259 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042123\
Data File : P0094490.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Apr 2023 12:38
Operator : YP/AJ
Sample : 02423-11DL 4X
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
SPE-1WC-23-04-18DL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 21 21:45:45 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041923.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 20 05:15:06 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.417 min
Delta R.T.: -0.001 min
Response: 19269698
Conc: 5.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPE-1WC-23-04-18DL




#1 Tetrachloro-m-xylene

R.T.: 3.619 min
Delta R.T.: -0.002 min
Response: 7040176
Conc: 5.45 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.258 min
Delta R.T.: 0.000 min
Response: 14606261
Conc: 6.83 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.637 min
Delta R.T.: -0.003 min
Response: 6098284
Conc: 6.12 ng/ml



#26 AR-1254-1


R.T.: 6.460 min
Delta R.T.: 0.004 min
Response: 51650693
Conc: 417.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPE-1WC-23-04-18DL




#26 AR-1254-1

R.T.: 5.504 min
Delta R.T.: 0.003 min
Response: 31986949
Conc: 457.65 ng/ml




#27 AR-1254-2

R.T.: 6.681 min
Delta R.T.: 0.004 min
Response: 94430057
Conc: 521.59 ng/ml



#27 AR-1254-2


R.T.: 5.652 min
Delta R.T.: 0.003 min
Response: 25353057
Conc: 404.90 ng/ml



#28 AR-1254-3


R.T.: 7.050 min
Delta R.T.: 0.002 min
Response: 90708305
Conc: 531.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPE-1WC-23-04-18DL




#28 AR-1254-3

R.T.: 6.056 min
Delta R.T.: 0.003 min
Response: 47929505
Conc: 518.44 ng/ml




#29 AR-1254-4

R.T.: 7.339 min
Delta R.T.: 0.003 min
Response: 37412472
Conc: 396.99 ng/ml



#29 AR-1254-4


R.T.: 6.286 min
Delta R.T.: 0.003 min
Response: 18905888
Conc: 446.20 ng/ml



#30 AR-1254-5

R.T.: 7.764 min
Delta R.T.: 0.003 min
Response: 90400322
Conc: 757.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
SPE-1WC-23-04-18DL



#30 AR-1254-5

R.T.: 6.708 min
Delta R.T.: 0.006 min
Response: 158837722
Conc: 2119.26 ng/ml