

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042221\
 Data File : P0077176.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Apr 2021 11:35
 Operator : DD\AJ
 Sample : M2091-13 10X
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEE0664-END-3-1

Manual Integrations
 APPROVED

Ankita
 4/22/2021 4:50:23 PM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:55:03 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 09:50:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.894	3.890	166133	107743	1.572	2.048 #
2) SA Decachlor...	10.863	9.131	172050	66380	1.657	1.759
Target Compounds						
26) L6 AR-1254-1	7.125	5.995	1277010	922949	310.544m	319.737m
27) L6 AR-1254-2	7.353	6.155	1942756	842726	300.168m	326.414m
28) L6 AR-1254-3	7.736	6.571	1766813	1177695	261.097m	301.013m
29) L6 AR-1254-4	8.032	6.811	1123156	610638	228.386m	268.762m
30) L6 AR-1254-5	8.460	7.235	1165204	928670	209.308m	256.130

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042221\
Data File : PO077176.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Apr 2021 11:35
Operator : DD\AJ
Sample : M2091-13 10X
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-3-1

Manual Integrations
APPROVED

Ankita
4/22/2021 4:50:23 PM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:55:03 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 09:50:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

