

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042224\
 Data File : P0103003.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2024 14:29
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:13:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon Apr 22 16:13:01 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.254	3.373	259.2E6	141.9E6	50.000	50.000
2) SA Decachlor...	9.776	8.227	148.2E6	87584606	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.397	4.421	75279859	39376491	500.000	500.000
4) L1 AR-1016-2	5.418	4.438	109.3E6	59700426	500.000	500.000
5) L1 AR-1016-3	5.479	4.610	69527532	31918499	500.000	500.000
6) L1 AR-1016-4	5.576	4.651	55619386	27007065	500.000	500.000
7) L1 AR-1016-5	5.866	4.858	54231010	34576087	500.000	500.000
31) L7 AR-1260-1	6.974	5.868	89717014	63651701	500.000	500.000
32) L7 AR-1260-2	7.229	6.054	99620474	77032984	500.000	500.000
33) L7 AR-1260-3	7.584	6.204	72302250	69272396	500.000	500.000
34) L7 AR-1260-4	7.808	6.667	78818256	55894356	500.000	500.000
35) L7 AR-1260-5	8.111	6.909	148.5E6	129.9E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042224\
Data File : PO103003.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2024 14:29
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:13:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon Apr 22 16:13:01 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.254 min
Delta R.T.: 0.000 min
Response: 259211333
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.373 min
Delta R.T.: 0.000 min
Response: 141903028
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.776 min
Delta R.T.: 0.000 min
Response: 148248950
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.227 min
Delta R.T.: 0.000 min
Response: 87584606
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.397 min
Delta R.T.: 0.000 min
Response: 75279859
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.421 min
Delta R.T.: 0.000 min
Response: 39376491
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.418 min
Delta R.T.: 0.000 min
Response: 109282307
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.438 min
Delta R.T.: 0.000 min
Response: 59700426
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.479 min
Delta R.T.: 0.000 min
Response: 69527532
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.610 min
Delta R.T.: 0.000 min
Response: 31918499
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.576 min
Delta R.T.: 0.000 min
Response: 55619386
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 4.651 min
Delta R.T.: 0.000 min
Response: 27007065
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.866 min
Delta R.T.: 0.000 min
Response: 54231010
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 4.858 min
Delta R.T.: 0.000 min
Response: 34576087
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.974 min
Delta R.T.: 0.000 min
Response: 89717014
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 5.868 min
Delta R.T.: 0.000 min
Response: 63651701
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.229 min
Delta R.T.: 0.000 min
Response: 99620474
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.054 min
Delta R.T.: 0.000 min
Response: 77032984
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.584 min
Delta R.T.: 0.000 min
Response: 72302250
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.204 min
Delta R.T.: 0.000 min
Response: 69272396
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.808 min
Delta R.T.: 0.000 min
Response: 78818256
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 6.667 min
Delta R.T.: 0.000 min
Response: 55894356
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.111 min
Delta R.T.: 0.000 min
Response: 148476198
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 6.909 min
Delta R.T.: 0.000 min
Response: 129941645
Conc: 500.00 ng/ml