

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042225\
 Data File : P0110595.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:34
 Operator : YP/AJ
 Sample : Q1846-01 10X
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 JEI294J-1-1

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/22/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:51:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	12260185	6871422	1.401	1.377
2) SA Decachlor...	8.731	8.685	8261248	2134699	1.046m	1.109m
Target Compounds						
26) L6 AR-1254-1	5.585	5.564	4730947	2614670	8.479m	8.672m
27) L6 AR-1254-2	5.731	5.710	2690683	1551803	5.540m	5.863m
28) L6 AR-1254-3	6.139	6.114	6480606	3813666	8.274m	9.453m
29) L6 AR-1254-4	6.369	6.343	4750303	2562498	9.742m	11.155m
30) L6 AR-1254-5	6.790	6.760	7018747	5151077	10.014	15.502m#

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042225\
Data File : PO110595.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:34
Operator : YP/AJ
Sample : Q1846-01 10X
Misc :
ALS Vial : 10 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
JEI294J-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:51:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 12260185
Conc: 1.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
JJEI294J-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.002 min
Response: 6871422
Conc: 1.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.731 min
Delta R.T.: 0.000 min
Response: 8261248
Conc: 1.05 ng/ml m



#2 Decachlorobiphenyl

R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 2134699
Conc: 1.11 ng/ml m




#26 AR-1254-1

R.T.: 5.585 min
Delta R.T.: -0.002 min
Response: 4730947
Conc: 8.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
JEI294J-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2025
Supervised By :mohammad ahmed 04/24/2025




#26 AR-1254-1

R.T.: 5.564 min
Delta R.T.: -0.002 min
Response: 2614670
Conc: 8.67 ng/ml




#27 AR-1254-2

R.T.: 5.731 min
Delta R.T.: -0.005 min
Response: 2690683
Conc: 5.54 ng/ml



#27 AR-1254-2

R.T.: 5.710 min
Delta R.T.: -0.004 min
Response: 1551803
Conc: 5.86 ng/ml




#28 AR-1254-3

R.T.: 6.139 min
Delta R.T.: -0.002 min
Response: 6480606
Conc: 8.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
JEI294J-1-1


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2025
Supervised By :mohammad ahmed 04/24/2025




#28 AR-1254-3

R.T.: 6.114 min
Delta R.T.: -0.003 min
Response: 3813666
Conc: 9.45 ng/ml m




#29 AR-1254-4

R.T.: 6.369 min
Delta R.T.: -0.002 min
Response: 4750303
Conc: 9.74 ng/ml m



#29 AR-1254-4

R.T.: 6.343 min
Delta R.T.: -0.002 min
Response: 2562498
Conc: 11.16 ng/ml m




#30 AR-1254-5

R.T.: 6.790 min
Delta R.T.: 0.000 min
Response: 7018747
Conc: 10.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
JEI294J-1-1

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/22/2025
Supervised By :mohammad ahmed 04/24/2025



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: -0.001 min
Response: 5151077
Conc: 15.50 ng/ml m