

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042321\
 Data File : P0077191.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 22 Apr 2021 18:36
 Operator : DD\AJ
 Sample : M2091-16 10X
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 HEE0664-END-4-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:36:57 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 09:50:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.896	3.890	157821	94130	1.493	1.789
2) SA Decachlor...	10.866	9.131	172034	64494	1.657	1.709
Target Compounds						
26) L6 AR-1254-1	7.126	5.997	1066567	840393	259.368	291.137
27) L6 AR-1254-2	7.357	6.156	1697174	765493	262.224	296.499
28) L6 AR-1254-3	7.738	6.572	1486431	1057622	219.663	270.323
29) L6 AR-1254-4	8.034	6.812	1000462	558137	203.437	245.654
30) L6 AR-1254-5	8.461	7.236	1110546	801204	199.489	220.974



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042321\
Data File : P0077191.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 22 Apr 2021 18:36
Operator : DD\AJ
Sample : M2091-16 10X
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-4-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:36:57 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 09:50:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.896 min
Delta R.T.: 0.000 min
Response: 157821
Conc: 1.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-4-1




#1 Tetrachloro-m-xylene

R.T.: 3.890 min
Delta R.T.: -0.002 min
Response: 94130
Conc: 1.79 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.866 min
Delta R.T.: 0.002 min
Response: 172034
Conc: 1.66 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.131 min
Delta R.T.: -0.001 min
Response: 64494
Conc: 1.71 ng/ml



#26 AR-1254-1


R.T.: 7.126 min
Delta R.T.: 0.002 min
Response: 1066567
Conc: 259.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-4-1




#26 AR-1254-1

R.T.: 5.997 min
Delta R.T.: -0.002 min
Response: 840393
Conc: 291.14 ng/ml




#27 AR-1254-2

R.T.: 7.357 min
Delta R.T.: 0.001 min
Response: 1697174
Conc: 262.22 ng/ml



#27 AR-1254-2


R.T.: 6.156 min
Delta R.T.: -0.002 min
Response: 765493
Conc: 296.50 ng/ml



#28 AR-1254-3


R.T.: 7.738 min
Delta R.T.: 0.002 min
Response: 1486431
Conc: 219.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-4-1




#28 AR-1254-3

R.T.: 6.572 min
Delta R.T.: -0.002 min
Response: 1057622
Conc: 270.32 ng/ml




#29 AR-1254-4

R.T.: 8.034 min
Delta R.T.: 0.000 min
Response: 1000462
Conc: 203.44 ng/ml



#29 AR-1254-4


R.T.: 6.812 min
Delta R.T.: -0.002 min
Response: 558137
Conc: 245.65 ng/ml



#30 AR-1254-5

R.T.: 8.461 min
Delta R.T.: 0.002 min
Response: 1110546
Conc: 199.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
HEE0664-END-4-1



#30 AR-1254-5

R.T.: 7.236 min
Delta R.T.: -0.001 min
Response: 801204
Conc: 220.97 ng/ml