

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042324\
 Data File : P0103087.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 00:51
 Operator : YP/AJ
 Sample : P2225-01MS
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 TR-06-042224MS

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/24/2024
 Supervised By :Ankita Jodhani 04/24/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 01:51:35 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.252	3.371	113.7E6	50184096	22.553	18.133
2)	SA Decachlor...	9.773	8.225	75434659	38533296	25.978	22.444
Target Compounds							
3)	L1 AR-1016-1	5.395	4.420	80845614	39183973	550.679	498.269
4)	L1 AR-1016-2	5.417	4.437	117.1E6	59618284	542.344	508.017
5)	L1 AR-1016-3	5.478	4.608	74330694	33478544	546.529	540.505
6)	L1 AR-1016-4	5.575	4.650	63070321	28490628	584.243	537.840
7)	L1 AR-1016-5	5.863	4.857	56669726	35873729	540.659m	526.122
31)	L7 AR-1260-1	6.974	5.866	98300454	62141030	554.839	488.717
32)	L7 AR-1260-2	7.227	6.053	111.5E6	75078769	562.250	488.847
33)	L7 AR-1260-3	7.582	6.202	67692944	66452481	478.527	476.869
34)	L7 AR-1260-4	7.806	6.666	81362173	47367868	523.388	430.620
35)	L7 AR-1260-5	8.110	6.907	191.7E6	120.2E6	642.651	468.544 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042324\
Data File : P0103087.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 00:51
Operator : YP/AJ
Sample : P2225-01MS
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
TR-06-042224MS

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/24/2024
Supervised By :Ankita Jodhani 04/24/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 01:51:35 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

