

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042424\
 Data File : P0103112.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 24 Apr 2024 10:18
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 24 12:45:01 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.252	3.371	266.4E6	130.4E6	52.814	47.104
2)	SA Decachlor...	9.773	8.224	163.8E6	92447642	56.392	53.846
Target Compounds							
16)	L4 AR-1242-1	5.395	4.420	61009429	29480268	500.208	444.691
17)	L4 AR-1242-2	5.416	4.437	87853092	44316593	490.541	454.287
18)	L4 AR-1242-3	5.477	4.608	54596045	24072814	482.867	449.102
19)	L4 AR-1242-4	5.574	4.689	43978045	24484687	492.610	443.968
20)	L4 AR-1242-5	6.299	5.200	40396978	30042233	483.372	474.697

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042424\
Data File : PO103112.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 24 Apr 2024 10:18
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 24 12:45:01 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm

