

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042524\
 Data File : P0103196.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2024 14:24
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 01:17:48 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.250	3.369	254.9E6	120.3E6	50.541	43.462
2) SA Decachlor...	9.766	8.220	163.8E6	87126803	56.392	50.747
Target Compounds						
16) L4 AR-1242-1	5.392	4.416	59638658	28924533	488.970	436.308
17) L4 AR-1242-2	5.413	4.434	88189224	44438136	492.418	455.533
18) L4 AR-1242-3	5.474	4.605	54015819	26255738	477.735	489.827
19) L4 AR-1242-4	5.572	4.686	43657864	25533631	489.023	462.988
20) L4 AR-1242-5	6.295	5.196	41363054	31310577	494.932	494.738



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042524\
Data File : PO103196.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2024 14:24
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 01:17:48 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.250 min
Delta R.T.: -0.004 min
Response: 254900997
Conc: 50.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.369 min
Delta R.T.: -0.004 min
Response: 120283155
Conc: 43.46 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.766 min
Delta R.T.: -0.010 min
Response: 163750575
Conc: 56.39 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.220 min
Delta R.T.: -0.007 min
Response: 87126803
Conc: 50.75 ng/ml



#16 AR-1242-1


R.T.: 5.392 min
Delta R.T.: -0.007 min
Response: 59638658
Conc: 488.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.416 min
Delta R.T.: -0.002 min
Response: 28924533
Conc: 436.31 ng/ml




#17 AR-1242-2

R.T.: 5.413 min
Delta R.T.: -0.007 min
Response: 88189224
Conc: 492.42 ng/ml



#17 AR-1242-2


R.T.: 4.434 min
Delta R.T.: -0.002 min
Response: 44438136
Conc: 455.53 ng/ml



#18 AR-1242-3


R.T.: 5.474 min
Delta R.T.: -0.007 min
Response: 54015819
Conc: 477.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.605 min
Delta R.T.: -0.002 min
Response: 26255738
Conc: 489.83 ng/ml




#19 AR-1242-4

R.T.: 5.572 min
Delta R.T.: -0.007 min
Response: 43657864
Conc: 489.02 ng/ml



#19 AR-1242-4


R.T.: 4.686 min
Delta R.T.: -0.002 min
Response: 25533631
Conc: 462.99 ng/ml



#20 AR-1242-5

R.T.: 6.295 min
Delta R.T.: -0.008 min
Response: 41363054
Conc: 494.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.196 min
Delta R.T.: -0.002 min
Response: 31310577
Conc: 494.74 ng/ml