

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042524\
 Data File : P0103198.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 25 Apr 2024 14:58
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 01:18:46 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.251	3.369	260.0E6	119.7E6	51.551	43.250
2)	SA Decachlor...	9.765	8.220	164.6E6	89671662	56.671	52.229
Target Compounds							
26)	L6 AR-1254-1	6.232	5.196	80082409	63180809	504.534	469.719
27)	L6 AR-1254-2	6.446	5.341	110.2E6	56811112	470.563	480.716
28)	L6 AR-1254-3	6.809	5.736	109.4E6	81132474	472.010	463.335
29)	L6 AR-1254-4	7.091	5.961	76974777	51701653	518.817	509.684
30)	L6 AR-1254-5	7.504	6.371	84295370	74130835	506.839	492.044



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042524\
Data File : PO103198.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 25 Apr 2024 14:58
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 01:18:46 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.251 min
Delta R.T.: -0.003 min
Response: 259996855
Conc: 51.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.369 min
Delta R.T.: -0.004 min
Response: 119695745
Conc: 43.25 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.765 min
Delta R.T.: -0.011 min
Response: 164558292
Conc: 56.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.220 min
Delta R.T.: -0.007 min
Response: 89671662
Conc: 52.23 ng/ml



#26 AR-1254-1


R.T.: 6.232 min
Delta R.T.: -0.005 min
Response: 80082409
Conc: 504.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.196 min
Delta R.T.: -0.004 min
Response: 63180809
Conc: 469.72 ng/ml




#27 AR-1254-2

R.T.: 6.446 min
Delta R.T.: -0.005 min
Response: 110212973
Conc: 470.56 ng/ml



#27 AR-1254-2


R.T.: 5.341 min
Delta R.T.: -0.005 min
Response: 56811112
Conc: 480.72 ng/ml



#28 AR-1254-3


R.T.: 6.809 min
Delta R.T.: -0.005 min
Response: 109367122
Conc: 472.01 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.736 min
Delta R.T.: -0.005 min
Response: 81132474
Conc: 463.34 ng/ml




#29 AR-1254-4

R.T.: 7.091 min
Delta R.T.: -0.005 min
Response: 76974777
Conc: 518.82 ng/ml



#29 AR-1254-4


R.T.: 5.961 min
Delta R.T.: -0.005 min
Response: 51701653
Conc: 509.68 ng/ml



#30 AR-1254-5

R.T.: 7.504 min
Delta R.T.: -0.006 min
Response: 84295370
Conc: 506.84 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.371 min
Delta R.T.: -0.005 min
Response: 74130835
Conc: 492.04 ng/ml