

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042621\
 Data File : P0077296.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 26 Apr 2021 12:42
 Operator : DD\AJ
 Sample : M2146-01
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 26 17:54:08 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 21 09:50:22 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.895	3.889	2068906	1071378	19.577	20.364
2)	SA Decachlor...	10.860	9.127	1849937	627064	17.818	16.618
Target Compounds							
7)	L1 AR-1016-5	0.000	5.600f	0	33483	N.D.	22.633 #
8)	L2 AR-1221-1	5.147	0.000	2611808	0	2302.784	N.D. #
12)	L3 AR-1232-2	5.946f	0.000	179188	0	161.047	N.D. #
15)	L3 AR-1232-5	0.000	5.600f	0	33483	N.D.	56.217 #
24)	L5 AR-1248-4	0.000	5.600f	0	33483	N.D.	17.842 #
30)	L6 AR-1254-5	0.000	7.230	0	37542	N.D.	10.354 #
36)	L8 AR-1262-1	0.000	7.230	0	37542	N.D.	20.103 #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042621\
Data File : P0077296.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 26 Apr 2021 12:42
Operator : DD\AJ
Sample : M2146-01
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 26 17:54:08 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042121.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 21 09:50:22 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.895 min
Delta R.T.: 0.000 min
Response: 2068906
Conc: 19.58 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.889 min
Delta R.T.: -0.004 min
Response: 1071378
Conc: 20.36 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.860 min
Delta R.T.: -0.004 min
Response: 1849937
Conc: 17.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.127 min
Delta R.T.: -0.006 min
Response: 627064
Conc: 16.62 ng/ml



#7 AR-1016-5


R.T.: 0.000 min
Exp R.T.: 6.729 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 5.600 min
Delta R.T.: -0.022 min
Response: 33483
Conc: 22.63 ng/ml




#8 AR-1221-1

R.T.: 5.147 min
Delta R.T.: -0.002 min
Response: 2611808
Conc: 2302.78 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T.: 4.150 min
Response: 0
Conc: N.D.



#12 AR-1232-2


R.T.: 5.946 min
Delta R.T.: 0.024 min
Response: 179188
Conc: 161.05 ng/ml

Instrument :
ECD_O
ClientSampleId :




#12 AR-1232-2

R.T.: 0.000 min
Exp R.T. : 5.159 min
Response: 0
Conc: N.D.




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 6.512 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.600 min
Delta R.T.: -0.023 min
Response: 33483
Conc: 56.22 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T.: 7.158 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :




#24 AR-1248-4

R.T.: 5.600 min
Delta R.T.: -0.022 min
Response: 33483
Conc: 17.84 ng/ml




#30 AR-1254-5

R.T.: 0.000 min
Exp R.T.: 8.460 min
Response: 0
Conc: N.D.



#30 AR-1254-5


R.T.: 7.230 min
Delta R.T.: -0.007 min
Response: 37542
Conc: 10.35 ng/ml



#36 AR-1262-1

R.T.: 0.000 min
Exp R.T.: 8.532 min
Response: 0
Conc: N.D.

Instrument :
ECD_O
ClientSampleId :



#36 AR-1262-1

R.T.: 7.230 min
Delta R.T.: -0.005 min
Response: 37542
Conc: 20.10 ng/ml