

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0042623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 27 01:54:40 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094560.D 1000 =P0094558.D 750 =P0094559.D
 250 =P0094561.D 50 =P0094562.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.486	3.442	3.488	3.536	3.422	3.475	E6	1.28
2) SA	Decachlorobiphenyl	2.403	2.262	2.352	2.446	2.348	2.362	E6	2.92
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1	8.227	7.823	8.102	9.095	8.521	8.354	E4	5.80
17) L4	AR-1242-2	1.263	1.195	1.218	1.301	1.308	1.257	E5	3.96
18) L4	AR-1242-3	8.114	7.560	7.813	8.603	8.628	8.143	E4	5.81
19) L4	AR-1242-4	6.193	5.864	6.039	6.494	6.117	6.141	E4	3.78
20) L4	AR-1242-5	6.498	6.020	6.252	6.628	6.779	6.435	E4	4.70
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1						0.000		-1.00
42) L9	AR-1268-2						0.000		-1.00
43) L9	AR-1268-3						0.000		-1.00
44) L9	AR-1268-4						0.000		-1.00
45) L9	AR-1268-5						0.000		-1.00

Signal #2 Calibration Files

500 =P0094560.D 1000 =P0094558.D 750 =P0094559.D
 250 =P0094561.D 50 =P0094562.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.367	1.335	1.356	1.407	1.160	1.325	E6	7.24
2) SA	Decachlorobiphenyl	1.119	1.030	1.073	1.194	1.230	1.129	E6	7.35
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0042623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 27 01:54:40 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094560.D 1000 =P0094558.D 750 =P0094559.D
 250 =P0094561.D 50 =P0094562.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1	3.327	3.093	3.190	3.608	3.342	3.312 E4	5.88
17) L4	AR-1242-2	4.963	4.639	4.833	5.273	4.939	4.929 E4	4.68
18) L4	AR-1242-3	2.672	2.486	2.575	2.798	2.478	2.602 E4	5.19
19) L4	AR-1242-4	3.016	2.739	2.871	3.216	2.959	2.960 E4	5.97
20) L4	AR-1242-5	3.175	2.935	3.066	3.382	3.240	3.160 E4	5.38
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1						0.000	-1.00
42) L9	AR-1268-2						0.000	-1.00
43) L9	AR-1268-3						0.000	-1.00
44) L9	AR-1268-4						0.000	-1.00
45) L9	AR-1268-5						0.000	-1.00

(#) = Out of Range