

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0042623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 27 02:21:44 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094576.D 1000 =P0094574.D 750 =P0094575.D
 250 =P0094577.D 50 =P0094578.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	3.402	3.281	3.354	3.496	3.060	3.319	E6	4.94
2) SA	Decachlorobiphenyl	4.175	3.885	3.996	4.397	4.052	4.101	E6	4.78
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00
6) L1	AR-1016-4						0.000		-1.00
7) L1	AR-1016-5						0.000		-1.00
8) L2	AR-1221-1						0.000		-1.00
9) L2	AR-1221-2						0.000		-1.00
10) L2	AR-1221-3						0.000		-1.00
11) L3	AR-1232-1						0.000		-1.00
12) L3	AR-1232-2						0.000		-1.00
13) L3	AR-1232-3						0.000		-1.00
14) L3	AR-1232-4						0.000		-1.00
15) L3	AR-1232-5						0.000		-1.00
16) L4	AR-1242-1						0.000		-1.00
17) L4	AR-1242-2						0.000		-1.00
18) L4	AR-1242-3						0.000		-1.00
19) L4	AR-1242-4						0.000		-1.00
20) L4	AR-1242-5						0.000		-1.00
21) L5	AR-1248-1						0.000		-1.00
22) L5	AR-1248-2						0.000		-1.00
23) L5	AR-1248-3						0.000		-1.00
24) L5	AR-1248-4						0.000		-1.00
25) L5	AR-1248-5						0.000		-1.00
26) L6	AR-1254-1						0.000		-1.00
27) L6	AR-1254-2						0.000		-1.00
28) L6	AR-1254-3						0.000		-1.00
29) L6	AR-1254-4						0.000		-1.00
30) L6	AR-1254-5						0.000		-1.00
31) L7	AR-1260-1						0.000		-1.00
32) L7	AR-1260-2						0.000		-1.00
33) L7	AR-1260-3						0.000		-1.00
34) L7	AR-1260-4						0.000		-1.00
35) L7	AR-1260-5						0.000		-1.00
36) L8	AR-1262-1						0.000		-1.00
37) L8	AR-1262-2						0.000		-1.00
38) L8	AR-1262-3						0.000		-1.00
39) L8	AR-1262-4						0.000		-1.00
40) L8	AR-1262-5						0.000		-1.00
41) L9	AR-1268-1	3.417	3.210	3.314	3.535	3.435	3.382	E5	3.67
42) L9	AR-1268-2	3.087	2.872	2.975	3.175	3.063	3.035	E5	3.80
43) L9	AR-1268-3	2.757	2.597	2.662	2.869	2.605	2.698	E5	4.25
44) L9	AR-1268-4	0.984	0.912	0.955	1.017	0.826	0.939	E5	7.86
45) L9	AR-1268-5	8.570	8.231	8.398	8.838	8.003	8.408	E5	3.79

Signal #2 Calibration Files

500 =P0094576.D 1000 =P0094574.D 750 =P0094575.D
 250 =P0094577.D 50 =P0094578.D

	Compound	500	1000	750	250	50	Avg		%RSD
1) SA	Tetrachloro-m-xylene	1.338	1.274	1.300	1.382	1.227	1.304	E6	4.54
2) SA	Decachlorobiphenyl	1.898	1.749	1.799	2.021	2.069	1.907	E6	7.22
3) L1	AR-1016-1						0.000		-1.00
4) L1	AR-1016-2						0.000		-1.00
5) L1	AR-1016-3						0.000		-1.00

Method Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\
 Method File : P0042623.M
 Title : GC EXTRACTABLES
 Last Update : Thu Apr 27 02:21:44 2023
 Response Via : Initial Calibration

Calibration Files

500 =P0094576.D 1000 =P0094574.D 750 =P0094575.D
 250 =P0094577.D 50 =P0094578.D

	Compound	500	1000	750	250	50	Avg	%RSD
6) L1	AR-1016-4						0.000	-1.00
7) L1	AR-1016-5						0.000	-1.00
8) L2	AR-1221-1						0.000	-1.00
9) L2	AR-1221-2						0.000	-1.00
10) L2	AR-1221-3						0.000	-1.00
11) L3	AR-1232-1						0.000	-1.00
12) L3	AR-1232-2						0.000	-1.00
13) L3	AR-1232-3						0.000	-1.00
14) L3	AR-1232-4						0.000	-1.00
15) L3	AR-1232-5						0.000	-1.00
16) L4	AR-1242-1						0.000	-1.00
17) L4	AR-1242-2						0.000	-1.00
18) L4	AR-1242-3						0.000	-1.00
19) L4	AR-1242-4						0.000	-1.00
20) L4	AR-1242-5						0.000	-1.00
21) L5	AR-1248-1						0.000	-1.00
22) L5	AR-1248-2						0.000	-1.00
23) L5	AR-1248-3						0.000	-1.00
24) L5	AR-1248-4						0.000	-1.00
25) L5	AR-1248-5						0.000	-1.00
26) L6	AR-1254-1						0.000	-1.00
27) L6	AR-1254-2						0.000	-1.00
28) L6	AR-1254-3						0.000	-1.00
29) L6	AR-1254-4						0.000	-1.00
30) L6	AR-1254-5						0.000	-1.00
31) L7	AR-1260-1						0.000	-1.00
32) L7	AR-1260-2						0.000	-1.00
33) L7	AR-1260-3						0.000	-1.00
34) L7	AR-1260-4						0.000	-1.00
35) L7	AR-1260-5						0.000	-1.00
36) L8	AR-1262-1						0.000	-1.00
37) L8	AR-1262-2						0.000	-1.00
38) L8	AR-1262-3						0.000	-1.00
39) L8	AR-1262-4						0.000	-1.00
40) L8	AR-1262-5						0.000	-1.00
41) L9	AR-1268-1	1.694	1.578	1.626	1.790	1.759	1.690 E5	5.25
42) L9	AR-1268-2	1.557	1.457	1.494	1.653	1.502	1.532 E5	4.97
43) L9	AR-1268-3	1.363	1.270	1.302	1.441	1.425	1.360 E5	5.48
44) L9	AR-1268-4	5.150	4.734	4.899	5.521	4.675	4.996 E4	6.94
45) L9	AR-1268-5	3.935	3.752	3.836	4.071	3.983	3.915 E5	3.18

(#) = Out of Range