

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042623\
 Data File : P0094581.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2023 00:09
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO042623AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 02:15:03 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 02:01:00 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.618	175.1E6	68260208	52.265	52.123
2) SA Decachlor...	10.253	8.633	122.3E6	56772370	51.356	50.306
Target Compounds						
21) L5 AR-1248-1	5.594	4.700	31967792	13190038	511.529	517.167
22) L5 AR-1248-2	5.871	4.936	53667749	21212249	512.912	512.190
23) L5 AR-1248-3	6.076	4.978	57153404	22151600	516.178	516.322
24) L5 AR-1248-4	6.484	5.149	52195978	25103968	506.952	512.163
25) L5 AR-1248-5	6.524	5.541	56012938	20490140	515.019	515.012



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042623\
Data File : PO094581.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2023 00:09
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO042623AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 02:15:03 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 02:01:00 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 175067451
Conc: 52.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042623AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: 0.000 min
Response: 68260208
Conc: 52.12 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.253 min
Delta R.T.: -0.001 min
Response: 122332988
Conc: 51.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: 0.000 min
Response: 56772370
Conc: 50.31 ng/ml



#21 AR-1248-1


R.T.: 5.594 min
Delta R.T.: 0.000 min
Response: 31967792
Conc: 511.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042623AR1248




#21 AR-1248-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 13190038
Conc: 517.17 ng/ml




#22 AR-1248-2

R.T.: 5.871 min
Delta R.T.: 0.000 min
Response: 53667749
Conc: 512.91 ng/ml



#22 AR-1248-2


R.T.: 4.936 min
Delta R.T.: -0.001 min
Response: 21212249
Conc: 512.19 ng/ml



#23 AR-1248-3


R.T.: 6.076 min
Delta R.T.: 0.000 min
Response: 57153404
Conc: 516.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042623AR1248




#23 AR-1248-3

R.T.: 4.978 min
Delta R.T.: 0.000 min
Response: 22151600
Conc: 516.32 ng/ml




#24 AR-1248-4

R.T.: 6.484 min
Delta R.T.: 0.000 min
Response: 52195978
Conc: 506.95 ng/ml



#24 AR-1248-4


R.T.: 5.149 min
Delta R.T.: 0.000 min
Response: 25103968
Conc: 512.16 ng/ml



#25 AR-1248-5

R.T.: 6.524 min
Delta R.T.: 0.000 min
Response: 56012938
Conc: 515.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042623AR1248



#25 AR-1248-5

R.T.: 5.541 min
Delta R.T.: 0.000 min
Response: 20490140
Conc: 515.01 ng/ml