

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042623\
 Data File : P0094584.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 27 Apr 2023 01:00
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 27 03:11:00 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 27 03:04:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.618	175.8E6	68981208	51.043	51.553
2) SA Decachlor...	10.252	8.633	122.5E6	56489349	50.488	50.187
Target Compounds						
3) L1 AR-1016-1	5.594	4.700	50145964	19774991	492.264	498.592
4) L1 AR-1016-2	5.616	4.718	76026565	30217138	502.468	511.148
5) L1 AR-1016-3	5.679	4.894	49295498	16198293	497.209	516.006
6) L1 AR-1016-4	5.777	4.936	37759674	14945910	507.309	516.867
7) L1 AR-1016-5	6.076	5.149	41896714	18829403	502.162	514.486
31) L7 AR-1260-1	7.215	6.184	73656435	35082816	496.775	498.334
32) L7 AR-1260-2	7.475	6.373	79311372	39353247	489.080	492.263
33) L7 AR-1260-3	7.839	6.526	65857698	37496112	503.756	498.600
34) L7 AR-1260-4	8.066	6.999	69950416	31316805	500.244	500.274
35) L7 AR-1260-5	8.393	7.244	119.0E6	62886801	486.274	498.948



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042623\
Data File : PO094584.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 27 Apr 2023 01:00
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 27 03:11:00 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 27 03:04:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.002 min
Response: 175822616
Conc: 51.04 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: 0.001 min
Response: 68981208
Conc: 51.55 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.252 min
Delta R.T.: 0.002 min
Response: 122527148
Conc: 50.49 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: 0.002 min
Response: 56489349
Conc: 50.19 ng/ml



#3 AR-1016-1


R.T.: 5.594 min
Delta R.T.: 0.001 min
Response: 50145964
Conc: 492.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.700 min
Delta R.T.: 0.001 min
Response: 19774991
Conc: 498.59 ng/ml




#4 AR-1016-2

R.T.: 5.616 min
Delta R.T.: 0.002 min
Response: 76026565
Conc: 502.47 ng/ml



#4 AR-1016-2


R.T.: 4.718 min
Delta R.T.: 0.001 min
Response: 30217138
Conc: 511.15 ng/ml



#5 AR-1016-3


R.T.: 5.679 min
Delta R.T.: 0.002 min
Response: 49295498
Conc: 497.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.894 min
Delta R.T.: 0.001 min
Response: 16198293
Conc: 516.01 ng/ml




#6 AR-1016-4

R.T.: 5.777 min
Delta R.T.: 0.001 min
Response: 37759674
Conc: 507.31 ng/ml



#6 AR-1016-4


R.T.: 4.936 min
Delta R.T.: 0.001 min
Response: 14945910
Conc: 516.87 ng/ml



#7 AR-1016-5


R.T.: 6.076 min
Delta R.T.: 0.002 min
Response: 41896714
Conc: 502.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.149 min
Delta R.T.: 0.002 min
Response: 18829403
Conc: 514.49 ng/ml




#31 AR-1260-1

R.T.: 7.215 min
Delta R.T.: 0.002 min
Response: 73656435
Conc: 496.77 ng/ml



#31 AR-1260-1


R.T.: 6.184 min
Delta R.T.: 0.001 min
Response: 35082816
Conc: 498.33 ng/ml



#32 AR-1260-2


R.T.: 7.475 min
Delta R.T.: 0.002 min
Response: 79311372
Conc: 489.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.373 min
Delta R.T.: 0.001 min
Response: 39353247
Conc: 492.26 ng/ml




#33 AR-1260-3

R.T.: 7.839 min
Delta R.T.: 0.003 min
Response: 65857698
Conc: 503.76 ng/ml



#33 AR-1260-3


R.T.: 6.526 min
Delta R.T.: 0.002 min
Response: 37496112
Conc: 498.60 ng/ml



#34 AR-1260-4


R.T.: 8.066 min
Delta R.T.: 0.002 min
Response: 69950416
Conc: 500.24 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.999 min
Delta R.T.: 0.001 min
Response: 31316805
Conc: 500.27 ng/ml



#35 AR-1260-5

R.T.: 8.393 min
Delta R.T.: 0.002 min
Response: 118967446
Conc: 486.27 ng/ml



#35 AR-1260-5

R.T.: 7.244 min
Delta R.T.: 0.001 min
Response: 62886801
Conc: 498.95 ng/ml