

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077392.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Apr 2021 23:44
 Operator : DD\AJ
 Sample : P0042821ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:56:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	109.736	119.310 E3	-8.7	107	0.00
2 SA Decachlorobiphenyl	109.961	107.256 E3	2.5	98	0.00
3 L1 AR-1016-1	3.880	3.845 E3	0.9	100	0.00
4 L1 AR-1016-2	6.453	6.833 E3	-5.9	105	0.00
5 L1 AR-1016-3	4.081	4.494 E3	-10.1	110	0.00
6 L1 AR-1016-4	3.146	3.416 E3	-8.6	107	0.00
7 L1 AR-1016-5	3.048	3.256 E3	-6.8	105	0.00
31 L7 AR-1260-1	6.182	6.312 E3	-2.1	104	0.00
32 L7 AR-1260-2	7.491	7.618 E3	-1.7	103	0.00
33 L7 AR-1260-3	5.006	5.255 E3	-5.0	106	0.00
34 L7 AR-1260-4	5.476	5.851 E3	-6.8	108	0.00
35 L7 AR-1260-5	10.848	11.224 E3	-3.5	104	0.00

Signal #2

1 SA Tetrachloro-m-xylene	55.823	57.378 E3	-2.8	102	0.00
2 SA Decachlorobiphenyl	34.019	32.639 E3	4.1	98	0.00
3 L1 AR-1016-1	1.538	1.384 E3	10.0	91	0.00
4 L1 AR-1016-2	3.312	3.298 E3	0.4	99	0.00
5 L1 AR-1016-3	1.445	1.479 E3	-2.4	99	0.00
6 L1 AR-1016-4	1.379	1.483 E3	-7.5	107	0.00
7 L1 AR-1016-5	1.596	1.483 E3	7.1	93	0.00
31 L7 AR-1260-1	3.147	3.158 E3	-0.3	102	0.00
32 L7 AR-1260-2	3.719	3.685 E3	0.9	100	0.00
33 L7 AR-1260-3	3.440	3.327 E3	3.3	98	0.00
34 L7 AR-1260-4	2.389	2.316 E3	3.1	98	0.00
35 L7 AR-1260-5	5.241	5.111 E3	2.5	97	0.00

Evaluate Continuing Calibration Report - Not Found

8 L2 AR-1221-1	1.179	0.000 E3	100.0#	0#	-5.15#
9 L2 AR-1221-2	868.715	0.000	100.0#	0#	-5.25#
10 L2 AR-1221-3	2.876	0.000 E3	100.0#	0#	-5.33#
11 L3 AR-1232-1	2.222	0.000 E3	100.0#	0#	-5.33#
12 L3 AR-1232-2	1.382	0.000 E3	100.0#	0#	-5.92#
13 L3 AR-1232-3	2.570	0.000 E3	100.0#	0#	-6.24#
14 L3 AR-1232-4	1.253	0.000 E3	100.0#	0#	-6.41#
15 L3 AR-1232-5	884.420	0.000	100.0#	0#	-6.51#
16 L4 AR-1242-1	2.867	0.000 E3	100.0#	0#	-6.22#
17 L4 AR-1242-2	4.937	0.000 E3	100.0#	0#	-6.24#
18 L4 AR-1242-3	3.185	0.000 E3	100.0#	0#	-6.31#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077392.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Apr 2021 23:44
 Operator : DD\AJ
 Sample : P0042821ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:56:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
19 L4 AR-1242-4	2.369	0.000 E3	100.0#	0#	-6.41#
20 L4 AR-1242-5	2.760	0.000 E3	100.0#	0#	-7.20#
21 L5 AR-1248-1	2.194	0.000 E3	100.0#	0#	-6.22#
22 L5 AR-1248-2	3.422	0.000 E3	100.0#	0#	-6.51#
23 L5 AR-1248-3	4.009	0.000 E3	100.0#	0#	-6.73#
24 L5 AR-1248-4	4.631	0.000 E3	100.0#	0#	-7.16#
25 L5 AR-1248-5	4.738	0.000 E3	100.0#	0#	-7.20#
26 L6 AR-1254-1	4.714	0.000 E3	100.0#	0#	-7.12#
27 L6 AR-1254-2	7.427	0.000 E3	100.0#	0#	-7.35#
28 L6 AR-1254-3	7.759	0.000 E3	100.0#	0#	-7.74#
29 L6 AR-1254-4	5.758	0.000 E3	100.0#	0#	-8.03#
30 L6 AR-1254-5	6.268	0.000 E3	100.0#	0#	-8.46#
36 L8 AR-1262-1	7.970	0.000 E3	100.0#	0#	-8.53#
37 L8 AR-1262-2	13.103	0.000 E3	100.0#	0#	-9.08#
38 L8 AR-1262-3	8.727	0.000 E3	100.0#	0#	-9.39#
39 L8 AR-1262-4	6.966	0.000 E3	100.0#	0#	-9.48#
40 L8 AR-1262-5	4.680	0.000 E3	100.0#	0#	-10.13#
41 L9 AR-1268-1	16.031	0.000 E3	100.0#	0#	-9.39#
42 L9 AR-1268-2	14.918	0.000 E3	100.0#	0#	-9.48#
43 L9 AR-1268-3	12.893	0.000 E3	100.0#	0#	-9.71#
44 L9 AR-1268-4	5.014	0.000 E3	100.0#	0#	-10.14#
45 L9 AR-1268-5	40.080	0.000 E3	100.0#	0#	-10.54#

Signal #2

8 L2 AR-1221-1	603.209	0.000	100.0#	0#	-4.15#
9 L2 AR-1221-2	438.026	0.000	100.0#	0#	-4.25#
10 L2 AR-1221-3	1.407	0.000 E3	100.0#	0#	-4.33#
11 L3 AR-1232-1	1.081	0.000 E3	100.0#	0#	-4.33#
12 L3 AR-1232-2	1.317	0.000 E3	100.0#	0#	-5.15#
13 L3 AR-1232-3	529.125	0.000	100.0#	0#	-5.34#
14 L3 AR-1232-4	545.557	0.000	100.0#	0#	-5.44#
15 L3 AR-1232-5	588.047	0.000	100.0#	0#	-5.62#
16 L4 AR-1242-1	1.122	0.000 E3	100.0#	0#	-5.13#
17 L4 AR-1242-2	2.475	0.000 E3	100.0#	0#	-5.15#
18 L4 AR-1242-3	1.100	0.000 E3	100.0#	0#	-5.34#
19 L4 AR-1242-4	1.229	0.000 E3	100.0#	0#	-5.43#
20 L4 AR-1242-5	1.535	0.000 E3	100.0#	0#	-6.00#
21 L5 AR-1248-1	874.334	0.000	100.0#	0#	-5.14#
22 L5 AR-1248-2	1.786	0.000 E3	100.0#	0#	-5.39#
23 L5 AR-1248-3	1.710	0.000 E3	100.0#	0#	-5.44#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077392.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 28 Apr 2021 23:44
 Operator : DD\AJ
 Sample : P0042821ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:56:50 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	2.017	0.000 E3	100.0#	0#	-5.62#
25 L5 AR-1248-5	2.284	0.000 E3	100.0#	0#	-6.04#
26 L6 AR-1254-1	3.090	0.000 E3	100.0#	0#	-5.99#
27 L6 AR-1254-2	2.772	0.000 E3	100.0#	0#	-6.15#
28 L6 AR-1254-3	4.076	0.000 E3	100.0#	0#	-6.57#
29 L6 AR-1254-4	2.284	0.000 E3	100.0#	0#	-6.81#
30 L6 AR-1254-5	3.667	0.000 E3	100.0#	0#	-7.23#
36 L8 AR-1262-1	1.855	0.000 E3	100.0#	0#	-7.23#
37 L8 AR-1262-2	5.905	0.000 E3	100.0#	0#	-7.78#
38 L8 AR-1262-3	2.290	0.000 E3	100.0#	0#	-8.06#
39 L8 AR-1262-4	4.102	0.000 E3	100.0#	0#	-8.13#
40 L8 AR-1262-5	1.698	0.000 E3	100.0#	0#	-8.62#
41 L9 AR-1268-1	6.522	0.000 E3	100.0#	0#	-8.06#
42 L9 AR-1268-2	5.833	0.000 E3	100.0#	0#	-8.13#
43 L9 AR-1268-3	4.837	0.000 E3	100.0#	0#	-8.33#
44 L9 AR-1268-4	1.924	0.000 E3	100.0#	0#	-8.63#
45 L9 AR-1268-5	12.433	0.000 E3	100.0#	0#	-8.90#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0