

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077395.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2021 00:36
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:58:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
1 SA Tetrachloro-m-xylene	109.736	115.765 E3	-5.5	104	0.00
2 SA Decachlorobiphenyl	109.961	103.590 E3	5.8	95	0.00
26 L6 AR-1254-1	4.714	4.704 E3	0.2	100	0.00
27 L6 AR-1254-2	7.427	7.409 E3	0.2	99	0.00
28 L6 AR-1254-3	7.759	7.628 E3	1.7	99	0.00
29 L6 AR-1254-4	5.758	5.662 E3	1.7	98	0.00
30 L6 AR-1254-5	6.268	6.252 E3	0.3	99	0.00

Signal #2

1 SA Tetrachloro-m-xylene	55.823	55.090 E3	1.3	98	0.00
2 SA Decachlorobiphenyl	34.019	31.197 E3	8.3	94	0.00
26 L6 AR-1254-1	3.090	3.035 E3	1.8	98	0.00
27 L6 AR-1254-2	2.772	2.687 E3	3.1	98	0.00
28 L6 AR-1254-3	4.076	3.909 E3	4.1	96	0.00
29 L6 AR-1254-4	2.284	2.125 E3	7.0	91	0.00
30 L6 AR-1254-5	3.667	3.631 E3	1.0	97	0.00

Evaluate Continuing Calibration Report - Not Found

3 L1 AR-1016-1	3.880	0.000 E3	100.0#	0#	-6.22#
4 L1 AR-1016-2	6.453	0.000 E3	100.0#	0#	-6.24#
5 L1 AR-1016-3	4.081	0.000 E3	100.0#	0#	-6.31#
6 L1 AR-1016-4	3.146	0.000 E3	100.0#	0#	-6.41#
7 L1 AR-1016-5	3.048	0.000 E3	100.0#	0#	-6.73#
8 L2 AR-1221-1	1.179	0.000 E3	100.0#	0#	-5.15#
9 L2 AR-1221-2	868.715	0.000	100.0#	0#	-5.25#
10 L2 AR-1221-3	2.876	0.000 E3	100.0#	0#	-5.33#
11 L3 AR-1232-1	2.222	0.000 E3	100.0#	0#	-5.33#
12 L3 AR-1232-2	1.382	0.000 E3	100.0#	0#	-5.92#
13 L3 AR-1232-3	2.570	0.000 E3	100.0#	0#	-6.24#
14 L3 AR-1232-4	1.253	0.000 E3	100.0#	0#	-6.41#
15 L3 AR-1232-5	884.420	0.000	100.0#	0#	-6.51#
16 L4 AR-1242-1	2.867	0.000 E3	100.0#	0#	-6.22#
17 L4 AR-1242-2	4.937	0.000 E3	100.0#	0#	-6.24#
18 L4 AR-1242-3	3.185	0.000 E3	100.0#	0#	-6.31#
19 L4 AR-1242-4	2.369	0.000 E3	100.0#	0#	-6.41#
20 L4 AR-1242-5	2.760	0.000 E3	100.0#	0#	-7.20#
21 L5 AR-1248-1	2.194	0.000 E3	100.0#	0#	-6.22#
22 L5 AR-1248-2	3.422	0.000 E3	100.0#	0#	-6.51#
23 L5 AR-1248-3	4.009	0.000 E3	100.0#	0#	-6.73#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077395.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2021 00:36
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:58:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	4.631	0.000 E3	100.0#	0#	-7.16#
25 L5 AR-1248-5	4.738	0.000 E3	100.0#	0#	-7.20#
31 L7 AR-1260-1	6.182	0.000 E3	100.0#	0#	-7.90#
32 L7 AR-1260-2	7.491	0.000 E3	100.0#	0#	-8.17#
33 L7 AR-1260-3	5.006	0.000 E3	100.0#	0#	-8.53#
34 L7 AR-1260-4	5.476	0.000 E3	100.0#	0#	-8.76#
35 L7 AR-1260-5	10.848	0.000 E3	100.0#	0#	-9.08#
36 L8 AR-1262-1	7.970	0.000 E3	100.0#	0#	-8.53#
37 L8 AR-1262-2	13.103	0.000 E3	100.0#	0#	-9.08#
38 L8 AR-1262-3	8.727	0.000 E3	100.0#	0#	-9.39#
39 L8 AR-1262-4	6.966	0.000 E3	100.0#	0#	-9.48#
40 L8 AR-1262-5	4.680	0.000 E3	100.0#	0#	-10.13#
41 L9 AR-1268-1	16.031	0.000 E3	100.0#	0#	-9.39#
42 L9 AR-1268-2	14.918	0.000 E3	100.0#	0#	-9.48#
43 L9 AR-1268-3	12.893	0.000 E3	100.0#	0#	-9.71#
44 L9 AR-1268-4	5.014	0.000 E3	100.0#	0#	-10.14#
45 L9 AR-1268-5	40.080	0.000 E3	100.0#	0#	-10.54#

Signal #2

3 L1 AR-1016-1	1.538	0.000 E3	100.0#	0#	-5.13#
4 L1 AR-1016-2	3.312	0.000 E3	100.0#	0#	-5.15#
5 L1 AR-1016-3	1.445	0.000 E3	100.0#	0#	-5.34#
6 L1 AR-1016-4	1.379	0.000 E3	100.0#	0#	-5.39#
7 L1 AR-1016-5	1.596	0.000 E3	100.0#	0#	-5.62#
8 L2 AR-1221-1	603.209	0.000	100.0#	0#	-4.15#
9 L2 AR-1221-2	438.026	0.000	100.0#	0#	-4.25#
10 L2 AR-1221-3	1.407	0.000 E3	100.0#	0#	-4.33#
11 L3 AR-1232-1	1.081	0.000 E3	100.0#	0#	-4.33#
12 L3 AR-1232-2	1.317	0.000 E3	100.0#	0#	-5.15#
13 L3 AR-1232-3	529.125	0.000	100.0#	0#	-5.34#
14 L3 AR-1232-4	545.557	0.000	100.0#	0#	-5.44#
15 L3 AR-1232-5	588.047	0.000	100.0#	0#	-5.62#
16 L4 AR-1242-1	1.122	0.000 E3	100.0#	0#	-5.13#
17 L4 AR-1242-2	2.475	0.000 E3	100.0#	0#	-5.15#
18 L4 AR-1242-3	1.100	0.000 E3	100.0#	0#	-5.34#
19 L4 AR-1242-4	1.229	0.000 E3	100.0#	0#	-5.43#
20 L4 AR-1242-5	1.535	0.000 E3	100.0#	0#	-6.00#
21 L5 AR-1248-1	874.334	0.000	100.0#	0#	-5.14#
22 L5 AR-1248-2	1.786	0.000 E3	100.0#	0#	-5.39#
23 L5 AR-1248-3	1.710	0.000 E3	100.0#	0#	-5.44#

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0042821\
 Data File : P0077395.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2021 00:36
 Operator : DD\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 38 Sample Multiplier: 1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 05:58:20 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 15% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(Min)
24 L5 AR-1248-4	2.017	0.000 E3	100.0#	0#	-5.62#
25 L5 AR-1248-5	2.284	0.000 E3	100.0#	0#	-6.04#
31 L7 AR-1260-1	3.147	0.000 E3	100.0#	0#	-6.70#
32 L7 AR-1260-2	3.719	0.000 E3	100.0#	0#	-6.90#
33 L7 AR-1260-3	3.440	0.000 E3	100.0#	0#	-7.05#
34 L7 AR-1260-4	2.389	0.000 E3	100.0#	0#	-7.53#
35 L7 AR-1260-5	5.241	0.000 E3	100.0#	0#	-7.78#
36 L8 AR-1262-1	1.855	0.000 E3	100.0#	0#	-7.23#
37 L8 AR-1262-2	5.905	0.000 E3	100.0#	0#	-7.78#
38 L8 AR-1262-3	2.290	0.000 E3	100.0#	0#	-8.06#
39 L8 AR-1262-4	4.102	0.000 E3	100.0#	0#	-8.13#
40 L8 AR-1262-5	1.698	0.000 E3	100.0#	0#	-8.62#
41 L9 AR-1268-1	6.522	0.000 E3	100.0#	0#	-8.06#
42 L9 AR-1268-2	5.833	0.000 E3	100.0#	0#	-8.13#
43 L9 AR-1268-3	4.837	0.000 E3	100.0#	0#	-8.33#
44 L9 AR-1268-4	1.924	0.000 E3	100.0#	0#	-8.63#
45 L9 AR-1268-5	12.433	0.000 E3	100.0#	0#	-8.90#

(#) = Out of Range

SPCC's out = 0 CCC's out = 0