

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042919\
 Data File : P0055701.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 30 Apr 2019 6:10
 Operator : SM/SJ
 Sample : P0042919ICV500
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO042919

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 06:23:20 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042919.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 30 05:21:50 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.283	3.596	1333265	1371618	47.402	47.019
2)	SA Decachlor...	9.885	8.550	2038963	1521620	47.740	47.899
Target Compounds							
3)	L1 AR-1016-1	5.445	4.669	671873	668416	463.665	458.720
4)	L1 AR-1016-2	5.467	4.688	946105	930039	469.194	467.290
5)	L1 AR-1016-3	5.528	4.863	627337	530456	467.746	479.444
6)	L1 AR-1016-4	5.626	4.902	517358	425636	471.644	475.587
7)	L1 AR-1016-5	5.918	5.114	556297	571844	470.750	462.428
31)	L7 AR-1260-1	7.036	6.139	1094558	1117460	476.413	465.125
32)	L7 AR-1260-2	7.293	6.325	1347753	1548293	470.249	466.634
33)	L7 AR-1260-3	7.650	6.479	1072310	1293084	477.750	463.599
34)	L7 AR-1260-4	7.874	6.947	1240964	1096312	482.519	468.148
35)	L7 AR-1260-5	8.184	7.187	2352095	2712892	485.180	484.079



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042919\
Data File : P0055701.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 30 Apr 2019 6:10
Operator : SM/SJ
Sample : P0042919ICV500
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVP0042919

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 06:23:20 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042919.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 30 05:21:50 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.283 min
Delta R.T.: -0.001 min
Response: 1333265
Conc: 47.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#1 Tetrachloro-m-xylene

R.T.: 3.596 min
Delta R.T.: 0.000 min
Response: 1371618
Conc: 47.02 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.885 min
Delta R.T.: -0.001 min
Response: 2038963
Conc: 47.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.550 min
Delta R.T.: 0.000 min
Response: 1521620
Conc: 47.90 ng/ml



#3 AR-1016-1


R.T.: 5.445 min
Delta R.T.: -0.001 min
Response: 671873
Conc: 463.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#3 AR-1016-1

R.T.: 4.669 min
Delta R.T.: 0.000 min
Response: 668416
Conc: 458.72 ng/ml




#4 AR-1016-2

R.T.: 5.467 min
Delta R.T.: 0.000 min
Response: 946105
Conc: 469.19 ng/ml



#4 AR-1016-2


R.T.: 4.688 min
Delta R.T.: 0.000 min
Response: 930039
Conc: 467.29 ng/ml



#5 AR-1016-3


R.T.: 5.528 min
Delta R.T.: 0.000 min
Response: 627337
Conc: 467.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#5 AR-1016-3

R.T.: 4.863 min
Delta R.T.: 0.000 min
Response: 530456
Conc: 479.44 ng/ml




#6 AR-1016-4

R.T.: 5.626 min
Delta R.T.: -0.002 min
Response: 517358
Conc: 471.64 ng/ml



#6 AR-1016-4


R.T.: 4.902 min
Delta R.T.: 0.000 min
Response: 425636
Conc: 475.59 ng/ml



#7 AR-1016-5


R.T.: 5.918 min
Delta R.T.: 0.000 min
Response: 556297
Conc: 470.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#7 AR-1016-5

R.T.: 5.114 min
Delta R.T.: 0.000 min
Response: 571844
Conc: 462.43 ng/ml




#31 AR-1260-1

R.T.: 7.036 min
Delta R.T.: 0.000 min
Response: 1094558
Conc: 476.41 ng/ml



#31 AR-1260-1


R.T.: 6.139 min
Delta R.T.: 0.000 min
Response: 1117460
Conc: 465.12 ng/ml



#32 AR-1260-2


R.T.: 7.293 min
Delta R.T.: 0.000 min
Response: 1347753
Conc: 470.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#32 AR-1260-2

R.T.: 6.325 min
Delta R.T.: 0.000 min
Response: 1548293
Conc: 466.63 ng/ml




#33 AR-1260-3

R.T.: 7.650 min
Delta R.T.: 0.000 min
Response: 1072310
Conc: 477.75 ng/ml



#33 AR-1260-3


R.T.: 6.479 min
Delta R.T.: 0.000 min
Response: 1293084
Conc: 463.60 ng/ml



#34 AR-1260-4


R.T.: 7.874 min
Delta R.T.: -0.002 min
Response: 1240964
Conc: 482.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO042919




#34 AR-1260-4

R.T.: 6.947 min
Delta R.T.: 0.000 min
Response: 1096312
Conc: 468.15 ng/ml



#35 AR-1260-5

R.T.: 8.184 min
Delta R.T.: 0.000 min
Response: 2352095
Conc: 485.18 ng/ml



#35 AR-1260-5

R.T.: 7.187 min
Delta R.T.: 0.000 min
Response: 2712892
Conc: 484.08 ng/ml