

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042921\
 Data File : P0077441.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 29 Apr 2021 23:52
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 01:28:33 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042821.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 29 05:54:04 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.899	3.895	2433472	1270099	22.176	22.752
2) SA Decachlor...	10.868	9.125	2188653	689099	19.904	20.256
Target Compounds						
9) L2 AR-1221-2	5.253	0.000	26492	0	30.495	N.D. #
24) L5 AR-1248-4	7.150	0.000	3601	0	0.778	N.D. #
28) L6 AR-1254-3	7.760f	0.000	29044	0	3.743	N.D. #



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042921\
Data File : P0077441.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 29 Apr 2021 23:52
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 01:28:33 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042821.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 29 05:54:04 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.899 min
Delta R.T.: 0.005 min
Response: 2433472
Conc: 22.18 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.895 min
Delta R.T.: 0.003 min
Response: 1270099
Conc: 22.75 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.868 min
Delta R.T.: 0.005 min
Response: 2188653
Conc: 19.90 ng/ml



#2 Decachlorobiphenyl


R.T.: 9.125 min
Delta R.T.: -0.001 min
Response: 689099
Conc: 20.26 ng/ml



#9 AR-1221-2


R.T.: 5.253 min
Delta R.T.: 0.007 min
Response: 26492
Conc: 30.50 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 4.247 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 7.150 min
Delta R.T.: -0.006 min
Response: 3601
Conc: 0.78 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.617 min
Response: 0
Conc: N.D.



#28 AR-1254-3

R.T.: 7.760 min
Delta R.T.: 0.024 min
Response: 29044
Conc: 3.74 ng/ml

Instrument :
ECD_O
ClientSampleId :



#28 AR-1254-3

R.T.: 0.000 min
Exp R.T. : 6.568 min
Response: 0
Conc: N.D.