

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042924\
 Data File : P0103296.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2024 19:44
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 07:45:51 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.251	3.369	266.8E6	122.3E6	52.898	44.189
2) SA Decachlor...	9.765	8.219	154.3E6	95449064	53.151	55.594
Target Compounds						
26) L6 AR-1254-1	6.232	5.195	84382678	63808907	531.626	474.388
27) L6 AR-1254-2	6.448	5.340	122.2E6	57239387	521.629	484.340
28) L6 AR-1254-3	6.809	5.735	123.0E6	85139597	530.636	486.219
29) L6 AR-1254-4	7.091	5.961	82228221	56946544	554.226	561.389
30) L6 AR-1254-5	7.505	6.369	87848183	76564596	528.201	508.199



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042924\
Data File : PO103296.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2024 19:44
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 07:45:51 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.251 min
Delta R.T.: -0.003 min
Response: 266791181
Conc: 52.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.369 min
Delta R.T.: -0.004 min
Response: 122295390
Conc: 44.19 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.765 min
Delta R.T.: -0.011 min
Response: 154338164
Conc: 53.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.219 min
Delta R.T.: -0.008 min
Response: 95449064
Conc: 55.59 ng/ml



#26 AR-1254-1


R.T.: 6.232 min
Delta R.T.: -0.004 min
Response: 84382678
Conc: 531.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.195 min
Delta R.T.: -0.005 min
Response: 63808907
Conc: 474.39 ng/ml




#27 AR-1254-2

R.T.: 6.448 min
Delta R.T.: -0.004 min
Response: 122173199
Conc: 521.63 ng/ml



#27 AR-1254-2


R.T.: 5.340 min
Delta R.T.: -0.005 min
Response: 57239387
Conc: 484.34 ng/ml



#28 AR-1254-3


R.T.: 6.809 min
Delta R.T.: -0.005 min
Response: 122950992
Conc: 530.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.735 min
Delta R.T.: -0.006 min
Response: 85139597
Conc: 486.22 ng/ml




#29 AR-1254-4

R.T.: 7.091 min
Delta R.T.: -0.005 min
Response: 82228221
Conc: 554.23 ng/ml



#29 AR-1254-4


R.T.: 5.961 min
Delta R.T.: -0.006 min
Response: 56946544
Conc: 561.39 ng/ml



#30 AR-1254-5

R.T.: 7.505 min
Delta R.T.: -0.005 min
Response: 87848183
Conc: 528.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.369 min
Delta R.T.: -0.007 min
Response: 76564596
Conc: 508.20 ng/ml