

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042924\
 Data File : P0103303.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2024 22:28
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 07:48:44 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042224.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 23 02:20:10 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.251	3.368	278.1E6	121.2E6	55.140	43.782
2) SA Decachlor...	9.765	8.218	156.1E6	94862463	53.746	55.252
Target Compounds						
26) L6 AR-1254-1	6.233	5.195	89450406	65652747	563.554	488.096
27) L6 AR-1254-2	6.447	5.340	124.3E6	59030828	530.896	499.499
28) L6 AR-1254-3	6.810	5.735	125.0E6	85694754	539.540	489.390
29) L6 AR-1254-4	7.092	5.961	83484226	57531011	562.691	567.150
30) L6 AR-1254-5	7.505	6.369	89277525	76923677	536.795	510.582



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042924\
Data File : PO103303.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2024 22:28
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 07:48:44 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042224.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 23 02:20:10 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.251 min
Delta R.T.: -0.003 min
Response: 278096333
Conc: 55.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.368 min
Delta R.T.: -0.005 min
Response: 121169856
Conc: 43.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.765 min
Delta R.T.: -0.011 min
Response: 156066281
Conc: 53.75 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.218 min
Delta R.T.: -0.009 min
Response: 94862463
Conc: 55.25 ng/ml



#26 AR-1254-1


R.T.: 6.233 min
Delta R.T.: -0.004 min
Response: 89450406
Conc: 563.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.195 min
Delta R.T.: -0.005 min
Response: 65652747
Conc: 488.10 ng/ml




#27 AR-1254-2

R.T.: 6.447 min
Delta R.T.: -0.004 min
Response: 124343715
Conc: 530.90 ng/ml



#27 AR-1254-2


R.T.: 5.340 min
Delta R.T.: -0.006 min
Response: 59030828
Conc: 499.50 ng/ml



#28 AR-1254-3


R.T.: 6.810 min
Delta R.T.: -0.004 min
Response: 125014147
Conc: 539.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 5.735 min
Delta R.T.: -0.006 min
Response: 85694754
Conc: 489.39 ng/ml




#29 AR-1254-4

R.T.: 7.092 min
Delta R.T.: -0.004 min
Response: 83484226
Conc: 562.69 ng/ml



#29 AR-1254-4


R.T.: 5.961 min
Delta R.T.: -0.006 min
Response: 57531011
Conc: 567.15 ng/ml



#30 AR-1254-5

R.T.: 7.505 min
Delta R.T.: -0.005 min
Response: 89277525
Conc: 536.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.369 min
Delta R.T.: -0.007 min
Response: 76923677
Conc: 510.58 ng/ml