

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050123\
 Data File : P0094667.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2023 02:43
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:25:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.617	160.5E6	63881260	46.596	47.742
2) SA Decachlor...	10.265	8.636	108.8E6	50325790	44.844	44.711
Target Compounds						
16) L4 AR-1242-1	5.595	4.699	36038406	15817766	431.407	477.618
17) L4 AR-1242-2	5.618	4.717	53857635	23659828	428.403	479.970
18) L4 AR-1242-3	5.681	4.893	34745447	12640793	426.668	485.832
19) L4 AR-1242-4	5.779	4.978	27054602	14052097	440.523	474.716
20) L4 AR-1242-5	6.525	5.501	29872366	15182508	464.193	480.535



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050123\
Data File : PO094667.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2023 02:43
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:25:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 160504005
Conc: 46.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: 0.000 min
Response: 63881260
Conc: 47.74 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.265 min
Delta R.T.: 0.013 min
Response: 108828938
Conc: 44.84 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.636 min
Delta R.T.: 0.005 min
Response: 50325790
Conc: 44.71 ng/ml



#16 AR-1242-1


R.T.: 5.595 min
Delta R.T.: 0.000 min
Response: 36038406
Conc: 431.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.699 min
Delta R.T.: -0.001 min
Response: 15817766
Conc: 477.62 ng/ml




#17 AR-1242-2

R.T.: 5.618 min
Delta R.T.: 0.001 min
Response: 53857635
Conc: 428.40 ng/ml



#17 AR-1242-2


R.T.: 4.717 min
Delta R.T.: -0.001 min
Response: 23659828
Conc: 479.97 ng/ml



#18 AR-1242-3


R.T.: 5.681 min
Delta R.T.: 0.000 min
Response: 34745447
Conc: 426.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.893 min
Delta R.T.: -0.001 min
Response: 12640793
Conc: 485.83 ng/ml




#19 AR-1242-4

R.T.: 5.779 min
Delta R.T.: 0.001 min
Response: 27054602
Conc: 440.52 ng/ml



#19 AR-1242-4


R.T.: 4.978 min
Delta R.T.: -0.001 min
Response: 14052097
Conc: 474.72 ng/ml



#20 AR-1242-5

R.T.: 6.525 min
Delta R.T.: 0.002 min
Response: 29872366
Conc: 464.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.501 min
Delta R.T.: 0.000 min
Response: 15182508
Conc: 480.53 ng/ml