

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050123\
 Data File : P0094668.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 02 May 2023 03:00
 Operator : YP/AJ
 Sample : AR1248CCC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 02 05:25:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.416	3.617	164.6E6	65521580	47.781	48.968
2) SA Decachlor...	10.264	8.636	112.9E6	52550059	46.509	46.687
Target Compounds						
21) L5 AR-1248-1	5.596	4.700	29342035	12952130	469.513	507.839
22) L5 AR-1248-2	5.872	4.936	47049229	20252834	449.657	489.024
23) L5 AR-1248-3	6.078	4.978	51212173	21193367	462.520	493.987
24) L5 AR-1248-4	6.487	5.149	49708325	24493779	482.790	499.714
25) L5 AR-1248-5	6.526	5.542	52851870	20420908	485.954	513.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050123\
Data File : PO094668.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 02 May 2023 03:00
Operator : YP/AJ
Sample : AR1248CCC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 02 05:25:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.416 min
Delta R.T.: 0.000 min
Response: 164586657
Conc: 47.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: 0.000 min
Response: 65521580
Conc: 48.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.264 min
Delta R.T.: 0.012 min
Response: 112869422
Conc: 46.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.636 min
Delta R.T.: 0.005 min
Response: 52550059
Conc: 46.69 ng/ml



#21 AR-1248-1


R.T.: 5.596 min
Delta R.T.: 0.001 min
Response: 29342035
Conc: 469.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#21 AR-1248-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 12952130
Conc: 507.84 ng/ml




#22 AR-1248-2

R.T.: 5.872 min
Delta R.T.: 0.001 min
Response: 47049229
Conc: 449.66 ng/ml



#22 AR-1248-2


R.T.: 4.936 min
Delta R.T.: 0.000 min
Response: 20252834
Conc: 489.02 ng/ml



#23 AR-1248-3


R.T.: 6.078 min
Delta R.T.: 0.002 min
Response: 51212173
Conc: 462.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500




#23 AR-1248-3

R.T.: 4.978 min
Delta R.T.: 0.000 min
Response: 21193367
Conc: 493.99 ng/ml




#24 AR-1248-4

R.T.: 6.487 min
Delta R.T.: 0.002 min
Response: 49708325
Conc: 482.79 ng/ml



#24 AR-1248-4


R.T.: 5.149 min
Delta R.T.: 0.000 min
Response: 24493779
Conc: 499.71 ng/ml



#25 AR-1248-5

R.T.: 6.526 min
Delta R.T.: 0.003 min
Response: 52851870
Conc: 485.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248CCC500



#25 AR-1248-5

R.T.: 5.542 min
Delta R.T.: 0.000 min
Response: 20420908
Conc: 513.27 ng/ml