

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050319\
 Data File : P0055812.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 May 2019 11:25
 Operator : SM/SJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 13:08:41 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 13:08:09 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.282	3.593	1540734	1402418	50.000	50.000
2) SA Decachlor...	9.881	8.544	1749757	1305163	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.492	3.804	217373	183570	500.000	500.000
9) L2 AR-1221-2	4.578	3.888	167669	138675	500.000	500.000
10) L2 AR-1221-3	4.654	3.964	502794	444251	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050319\
Data File : P0055812.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 May 2019 11:25
Operator : SM/SJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 13:08:41 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 13:08:09 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.282 min
Delta R.T.: 0.000 min
Response: 1540734
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: 0.000 min
Response: 1402418
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.881 min
Delta R.T.: 0.000 min
Response: 1749757
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.544 min
Delta R.T.: 0.000 min
Response: 1305163
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.492 min
Delta R.T.: 0.000 min
Response: 217373
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.804 min
Delta R.T.: 0.000 min
Response: 183570
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.578 min
Delta R.T.: 0.000 min
Response: 167669
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.888 min
Delta R.T.: 0.000 min
Response: 138675
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.654 min
Delta R.T.: 0.000 min
Response: 502794
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 3.964 min
Delta R.T.: 0.000 min
Response: 444251
Conc: 500.00 ng/ml