

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050319\
 Data File : P0055815.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 03 May 2019 12:15
 Operator : SM/SJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 13:16:55 2019
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050319.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri May 03 13:16:01 2019
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.278	3.593	1529701	1439327	50.000	50.000
2) SA Decachlor...	9.877	8.543	1834152	1375100	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.439	4.664	452560	419264	500.000	500.000
22) L5 AR-1248-2	5.710	4.898	666809	549382	500.000	500.000
23) L5 AR-1248-3	5.912	4.938	738365	567573	500.000	500.000
24) L5 AR-1248-4	6.314	5.109	871867	702088	500.000	500.000
25) L5 AR-1248-5	6.352	5.498	835164	748858	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050319\
Data File : PO055815.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 03 May 2019 12:15
Operator : SM/SJ
Sample : AR1248ICC500
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 13:16:55 2019
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050319.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri May 03 13:16:01 2019
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.278 min
Delta R.T.: 0.000 min
Response: 1529701
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.593 min
Delta R.T.: 0.000 min
Response: 1439327
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.877 min
Delta R.T.: 0.000 min
Response: 1834152
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.543 min
Delta R.T.: 0.000 min
Response: 1375100
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.439 min
Delta R.T.: 0.000 min
Response: 452560
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.664 min
Delta R.T.: 0.000 min
Response: 419264
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.710 min
Delta R.T.: 0.000 min
Response: 666809
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 549382
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 5.912 min
Delta R.T.: 0.000 min
Response: 738365
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 4.938 min
Delta R.T.: 0.000 min
Response: 567573
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.314 min
Delta R.T.: 0.000 min
Response: 871867
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.109 min
Delta R.T.: 0.000 min
Response: 702088
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.352 min
Delta R.T.: 0.000 min
Response: 835164
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.498 min
Delta R.T.: 0.000 min
Response: 748858
Conc: 500.00 ng/ml