

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050323\
 Data File : P0094683.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 03 May 2023 11:59
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 03 13:41:42 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.424	3.617	173.9E6	66629255	50.473	49.795
2)	SA Decachlor...	10.278	8.639	123.5E6	53699099	50.891	47.708
Target Compounds							
16)	L4 AR-1242-1	5.604	4.700	43490569	16594523	520.615	501.073
17)	L4 AR-1242-2	5.627	4.717	62917280	24553652	500.467	498.103
18)	L4 AR-1242-3	5.690	4.894	40517747	13240223	497.551	508.871
19)	L4 AR-1242-4	5.788	4.979	31340917	14720248	510.315	497.288
20)	L4 AR-1242-5	6.535	5.502	31685615	16458895	492.370	520.933



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050323\
Data File : PO094683.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 03 May 2023 11:59
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 03 13:41:42 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.424 min
Delta R.T.: 0.008 min
Response: 173860838
Conc: 50.47 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: 0.000 min
Response: 66629255
Conc: 49.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.278 min
Delta R.T.: 0.026 min
Response: 123504160
Conc: 50.89 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.639 min
Delta R.T.: 0.008 min
Response: 53699099
Conc: 47.71 ng/ml



#16 AR-1242-1


R.T.: 5.604 min
Delta R.T.: 0.010 min
Response: 43490569
Conc: 520.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.700 min
Delta R.T.: 0.000 min
Response: 16594523
Conc: 501.07 ng/ml




#17 AR-1242-2

R.T.: 5.627 min
Delta R.T.: 0.011 min
Response: 62917280
Conc: 500.47 ng/ml



#17 AR-1242-2


R.T.: 4.717 min
Delta R.T.: -0.001 min
Response: 24553652
Conc: 498.10 ng/ml



#18 AR-1242-3


R.T.: 5.690 min
Delta R.T.: 0.010 min
Response: 40517747
Conc: 497.55 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 13240223
Conc: 508.87 ng/ml




#19 AR-1242-4

R.T.: 5.788 min
Delta R.T.: 0.010 min
Response: 31340917
Conc: 510.32 ng/ml



#19 AR-1242-4


R.T.: 4.979 min
Delta R.T.: 0.000 min
Response: 14720248
Conc: 497.29 ng/ml



#20 AR-1242-5

R.T.: 6.535 min
Delta R.T.: 0.012 min
Response: 31685615
Conc: 492.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.502 min
Delta R.T.: 0.000 min
Response: 16458895
Conc: 520.93 ng/ml