

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050421\
Data File : P0077516.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 May 2021 18:11
Operator : DD\AJ
Sample : M2263-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 01:42:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042821.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 29 05:54:04 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.893	3.888	2297076	1164699	20.933	20.864
2) SA Decachlor...	10.858	9.120	2362989	698785	21.489	20.541

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050421\
Data File : P0077516.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 04 May 2021 18:11
Operator : DD\AJ
Sample : M2263-01
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
S-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 01:42:26 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042821.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 29 05:54:04 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.893 min
Delta R.T.: -0.001 min
Response: 2297076
Conc: 20.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
S-1




#1 Tetrachloro-m-xylene

R.T.: 3.888 min
Delta R.T.: -0.005 min
Response: 1164699
Conc: 20.86 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.858 min
Delta R.T.: -0.005 min
Response: 2362989
Conc: 21.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.120 min
Delta R.T.: -0.006 min
Response: 698785
Conc: 20.54 ng/ml