

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
 Data File : P0094731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 13:41
 Operator : YP/AJ
 Sample : 02603-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 18:22:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.415	3.616	35191156	14261316	10.216	10.658
2)	SA Decachlor...	10.256	8.629	41673886	17949224	17.172	15.947
Target Compounds							
3)	L1 AR-1016-1	5.579	0.000	258410	0	2.537	N.D. #
4)	L1 AR-1016-2	5.620	0.000	236302	0	1.562	N.D. #
5)	L1 AR-1016-3	5.682	0.000	393027	0	3.964	N.D. #
6)	L1 AR-1016-4	5.787	0.000	10263	0	0.138	N.D. #
7)	L1 AR-1016-5	6.079	0.000	4753071	0	56.969	N.D. #
8)	L2 AR-1221-1	4.623	0.000	729042	0	16.435	N.D. #
9)	L2 AR-1221-2	4.722	0.000	631304	0	19.403	N.D. #
10)	L2 AR-1221-3	4.794	0.000	94365	0	0.944	N.D. #
11)	L3 AR-1232-1	4.794	0.000	94365	0	1.116	N.D. #
12)	L3 AR-1232-2	5.323	0.000	73546	0	1.715	N.D. #
13)	L3 AR-1232-3	5.620	0.000	236302	0	3.253	N.D. #
14)	L3 AR-1232-4	5.787	0.000	10263	0	0.292	N.D. #
15)	L3 AR-1232-5	5.870	0.000	597204	0	18.113	N.D. #
16)	L4 AR-1242-1	5.579	0.000	258410	0	3.093	N.D. #
17)	L4 AR-1242-2	5.620	0.000	236302	0	1.880	N.D. #
18)	L4 AR-1242-3	5.682	0.000	393027	0	4.826	N.D. #
19)	L4 AR-1242-4	5.787	0.000	10263	0	0.167	N.D. #
20)	L4 AR-1242-5	6.524	5.498	711183	529902	11.051	16.772 #
21)	L5 AR-1248-1	5.579	0.000	258410	0	4.135	N.D. #
22)	L5 AR-1248-2	5.870	0.000	597204	0	5.708	N.D. #
23)	L5 AR-1248-3	6.079	0.000	4753071	0	42.927	N.D. #
24)	L5 AR-1248-4	6.483	0.000	584794	0	5.680	N.D. #
25)	L5 AR-1248-5	6.524	5.498f	711183	529902	6.539	13.319 #
26)	L6 AR-1254-1	6.457	5.498	1230579	529902	9.713	7.310
27)	L6 AR-1254-2	6.678	5.646	2255547	886541	11.801	13.587
28)	L6 AR-1254-3	7.048	6.050	4036363	2754693	21.400	27.741 #
29)	L6 AR-1254-4	7.338	6.280	2180840	837060	19.255	16.959
30)	L6 AR-1254-5	7.762	6.698	8633122	4037634	61.644	47.992
31)	L7 AR-1260-1	7.214	6.180	3092884	1642640	20.860	23.333
32)	L7 AR-1260-2	7.475	6.370	5098408	2132370	31.440	26.673
33)	L7 AR-1260-3	7.838	6.522	2453486	1717322	18.767	22.836
34)	L7 AR-1260-4	8.055	6.996	5144454	703241	36.790	11.234 #
35)	L7 AR-1260-5	8.396	7.240	4135820	2773701	16.905	22.007 #
36)	L8 AR-1262-1	7.838	6.789	2453486	505616	14.608	13.491
37)	L8 AR-1262-2	8.396	6.996	4135820	703241	16.450	9.426 #
38)	L8 AR-1262-3	8.748	7.525	26051543	434742	144.248	7.474 #
39)	L8 AR-1262-4	8.748f	7.587	26051543	1497302	280.567	14.783 #
40)	L8 AR-1262-5	9.480	8.086	1510740	383711	17.640	8.816 #
41)	L9 AR-1268-1	8.748f	7.525	26051543	434742	77.025	2.573 #

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
 Data File : P0094731.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 13:41
 Operator : YP/AJ
 Sample : 02603-01
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 18:22:58 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml
42) L9	AR-1268-2	8.865f	7.587	41216	1497302	0.136	9.771 #
43) L9	AR-1268-3	9.040	0.000	72539	0	0.269	N.D. #
44) L9	AR-1268-4	9.480	8.086	1510740	383711	16.092	7.681 #
45) L9	AR-1268-5	9.908	8.378	686255	291645	0.816	0.745



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050423\
Data File : PO094731.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 13:41
Operator : YP/AJ
Sample : 02603-01
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 18:22:58 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: -0.001 min
Response: 35191156
Conc: 10.22 ng/ml

Instrument :
ECD_O
ClientSampleId :




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.002 min
Response: 14261316
Conc: 10.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.256 min
Delta R.T.: 0.004 min
Response: 41673886
Conc: 17.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.629 min
Delta R.T.: -0.002 min
Response: 17949224
Conc: 15.95 ng/ml



#3 AR-1016-1


R.T.: 5.579 min
Delta R.T.: -0.015 min
Response: 258410
Conc: 2.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#3 AR-1016-1

R.T.: 0.000 min
Exp R.T. : 4.700 min
Response: 0
Conc: N.D.




#4 AR-1016-2

R.T.: 5.620 min
Delta R.T.: 0.004 min
Response: 236302
Conc: 1.56 ng/ml



#4 AR-1016-2


R.T.: 0.000 min
Exp R.T. : 4.717 min
Response: 0
Conc: N.D.



#5 AR-1016-3


R.T.: 5.682 min
Delta R.T.: 0.003 min
Response: 393027
Conc: 3.96 ng/ml

Instrument :
ECD_O
ClientSampleId :




#5 AR-1016-3

R.T.: 0.000 min
Exp R.T. : 4.894 min
Response: 0
Conc: N.D.




#6 AR-1016-4

R.T.: 5.787 min
Delta R.T.: 0.009 min
Response: 10263
Conc: 0.14 ng/ml



#6 AR-1016-4


R.T.: 0.000 min
Exp R.T. : 4.936 min
Response: 0
Conc: N.D.



#7 AR-1016-5


R.T.: 6.079 min
Delta R.T.: 0.004 min
Response: 4753071
Conc: 56.97 ng/ml

Instrument :
ECD_O
ClientSampleId :




#7 AR-1016-5

R.T.: 0.000 min
Exp R.T. : 5.149 min
Response: 0
Conc: N.D.




#8 AR-1221-1

R.T.: 4.623 min
Delta R.T.: 0.000 min
Response: 729042
Conc: 16.44 ng/ml



#8 AR-1221-1


R.T.: 0.000 min
Exp R.T. : 3.829 min
Response: 0
Conc: N.D.



#9 AR-1221-2


R.T.: 4.722 min
Delta R.T.: 0.012 min
Response: 631304
Conc: 19.40 ng/ml

Instrument :
ECD_O
ClientSampleId :




#9 AR-1221-2

R.T.: 0.000 min
Exp R.T. : 3.914 min
Response: 0
Conc: N.D.




#10 AR-1221-3

R.T.: 4.794 min
Delta R.T.: 0.007 min
Response: 94365
Conc: 0.94 ng/ml



#10 AR-1221-3


R.T.: 0.000 min
Exp R.T. : 3.990 min
Response: 0
Conc: N.D.



#11 AR-1232-1


R.T.: 4.794 min
Delta R.T.: 0.005 min
Response: 94365
Conc: 1.12 ng/ml

Instrument :
ECD_O
ClientSampleId :




#11 AR-1232-1

R.T.: 0.000 min
Exp R.T. : 3.991 min
Response: 0
Conc: N.D.




#12 AR-1232-2

R.T.: 5.323 min
Delta R.T.: 0.000 min
Response: 73546
Conc: 1.72 ng/ml



#12 AR-1232-2


R.T.: 0.000 min
Exp R.T. : 4.716 min
Response: 0
Conc: N.D.



#13 AR-1232-3


R.T.: 5.620 min
Delta R.T.: 0.005 min
Response: 236302
Conc: 3.25 ng/ml

Instrument :
ECD_O
ClientSampleId :




#13 AR-1232-3

R.T.: 0.000 min
Exp R.T. : 4.893 min
Response: 0
Conc: N.D.




#14 AR-1232-4

R.T.: 5.787 min
Delta R.T.: 0.010 min
Response: 10263
Conc: 0.29 ng/ml



#14 AR-1232-4


R.T.: 0.000 min
Exp R.T. : 4.977 min
Response: 0
Conc: N.D.



#15 AR-1232-5


R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 597204
Conc: 18.11 ng/ml

Instrument :
ECD_O
ClientSampleId :




#15 AR-1232-5

R.T.: 0.000 min
Exp R.T. : 5.148 min
Response: 0
Conc: N.D.




#16 AR-1242-1

R.T.: 5.579 min
Delta R.T.: -0.015 min
Response: 258410
Conc: 3.09 ng/ml



#16 AR-1242-1


R.T.: 0.000 min
Exp R.T. : 4.700 min
Response: 0
Conc: N.D.



#17 AR-1242-2


R.T.: 5.620 min
Delta R.T.: 0.004 min
Response: 236302
Conc: 1.88 ng/ml

Instrument :
ECD_O
ClientSampleId :




#17 AR-1242-2

R.T.: 0.000 min
Exp R.T. : 4.718 min
Response: 0
Conc: N.D.




#18 AR-1242-3

R.T.: 5.682 min
Delta R.T.: 0.002 min
Response: 393027
Conc: 4.83 ng/ml



#18 AR-1242-3


R.T.: 0.000 min
Exp R.T. : 4.895 min
Response: 0
Conc: N.D.



#19 AR-1242-4


R.T.: 5.787 min
Delta R.T.: 0.009 min
Response: 10263
Conc: 0.17 ng/ml

Instrument :
ECD_O
ClientSampleId :




#19 AR-1242-4

R.T.: 0.000 min
Exp R.T. : 4.979 min
Response: 0
Conc: N.D.




#20 AR-1242-5

R.T.: 6.524 min
Delta R.T.: 0.001 min
Response: 711183
Conc: 11.05 ng/ml



#20 AR-1242-5


R.T.: 5.498 min
Delta R.T.: -0.004 min
Response: 529902
Conc: 16.77 ng/ml



#21 AR-1248-1


R.T.: 5.579 min
Delta R.T.: -0.015 min
Response: 258410
Conc: 4.13 ng/ml

Instrument :
ECD_O
ClientSampleId :




#21 AR-1248-1

R.T.: 0.000 min
Exp R.T. : 4.700 min
Response: 0
Conc: N.D.




#22 AR-1248-2

R.T.: 5.870 min
Delta R.T.: 0.000 min
Response: 597204
Conc: 5.71 ng/ml



#22 AR-1248-2


R.T.: 0.000 min
Exp R.T. : 4.937 min
Response: 0
Conc: N.D.



#23 AR-1248-3


R.T.: 6.079 min
Delta R.T.: 0.003 min
Response: 4753071
Conc: 42.93 ng/ml

Instrument :
ECD_O
ClientSampleId :




#23 AR-1248-3

R.T.: 0.000 min
Exp R.T. : 4.978 min
Response: 0
Conc: N.D.




#24 AR-1248-4

R.T.: 6.483 min
Delta R.T.: -0.002 min
Response: 584794
Conc: 5.68 ng/ml



#24 AR-1248-4


R.T.: 0.000 min
Exp R.T. : 5.150 min
Response: 0
Conc: N.D.



#25 AR-1248-5


R.T.: 6.524 min
Delta R.T.: 0.001 min
Response: 711183
Conc: 6.54 ng/ml

Instrument :
ECD_O
ClientSampleId :




#25 AR-1248-5

R.T.: 5.498 min
Delta R.T.: -0.044 min
Response: 529902
Conc: 13.32 ng/ml




#26 AR-1254-1

R.T.: 6.457 min
Delta R.T.: -0.006 min
Response: 1230579
Conc: 9.71 ng/ml



#26 AR-1254-1


R.T.: 5.498 min
Delta R.T.: -0.006 min
Response: 529902
Conc: 7.31 ng/ml



#27 AR-1254-2


R.T.: 6.678 min
Delta R.T.: -0.007 min
Response: 2255547
Conc: 11.80 ng/ml

Instrument :
ECD_O
ClientSampleId :




#27 AR-1254-2

R.T.: 5.646 min
Delta R.T.: -0.006 min
Response: 886541
Conc: 13.59 ng/ml




#28 AR-1254-3

R.T.: 7.048 min
Delta R.T.: -0.008 min
Response: 4036363
Conc: 21.40 ng/ml



#28 AR-1254-3


R.T.: 6.050 min
Delta R.T.: -0.006 min
Response: 2754693
Conc: 27.74 ng/ml



#29 AR-1254-4


R.T.: 7.338 min
Delta R.T.: -0.007 min
Response: 2180840
Conc: 19.26 ng/ml

Instrument :
ECD_O
ClientSampleId :




#29 AR-1254-4

R.T.: 6.280 min
Delta R.T.: -0.006 min
Response: 837060
Conc: 16.96 ng/ml




#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: -0.008 min
Response: 8633122
Conc: 61.64 ng/ml



#30 AR-1254-5


R.T.: 6.698 min
Delta R.T.: -0.007 min
Response: 4037634
Conc: 47.99 ng/ml



#31 AR-1260-1


R.T.: 7.214 min
Delta R.T.: 0.000 min
Response: 3092884
Conc: 20.86 ng/ml

Instrument :
ECD_O
ClientSampleId :




#31 AR-1260-1

R.T.: 6.180 min
Delta R.T.: -0.003 min
Response: 1642640
Conc: 23.33 ng/ml




#32 AR-1260-2

R.T.: 7.475 min
Delta R.T.: 0.000 min
Response: 5098408
Conc: 31.44 ng/ml



#32 AR-1260-2


R.T.: 6.370 min
Delta R.T.: -0.003 min
Response: 2132370
Conc: 26.67 ng/ml



#33 AR-1260-3


R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 2453486
Conc: 18.77 ng/ml

Instrument :
ECD_O
ClientSampleId :




#33 AR-1260-3

R.T.: 6.522 min
Delta R.T.: -0.004 min
Response: 1717322
Conc: 22.84 ng/ml




#34 AR-1260-4

R.T.: 8.055 min
Delta R.T.: -0.012 min
Response: 5144454
Conc: 36.79 ng/ml



#34 AR-1260-4


R.T.: 6.996 min
Delta R.T.: -0.003 min
Response: 703241
Conc: 11.23 ng/ml



#35 AR-1260-5


R.T.: 8.396 min
Delta R.T.: 0.002 min
Response: 4135820
Conc: 16.90 ng/ml

Instrument :
ECD_O
ClientSampleId :




#35 AR-1260-5

R.T.: 7.240 min
Delta R.T.: -0.003 min
Response: 2773701
Conc: 22.01 ng/ml




#36 AR-1262-1

R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 2453486
Conc: 14.61 ng/ml



#36 AR-1262-1


R.T.: 6.789 min
Delta R.T.: -0.004 min
Response: 505616
Conc: 13.49 ng/ml



#37 AR-1262-2


R.T.: 8.396 min
Delta R.T.: 0.000 min
Response: 4135820
Conc: 16.45 ng/ml

Instrument :
ECD_O
ClientSampleId :




#37 AR-1262-2

R.T.: 6.996 min
Delta R.T.: -0.004 min
Response: 703241
Conc: 9.43 ng/ml




#38 AR-1262-3

R.T.: 8.748 min
Delta R.T.: 0.035 min
Response: 26051543
Conc: 144.25 ng/ml



#38 AR-1262-3


R.T.: 7.525 min
Delta R.T.: -0.003 min
Response: 434742
Conc: 7.47 ng/ml



#39 AR-1262-4


R.T.: 8.748 min
Delta R.T.: -0.057 min
Response: 26051543
Conc: 280.57 ng/ml

Instrument :
ECD_O
ClientSampleId :




#39 AR-1262-4

R.T.: 7.587 min
Delta R.T.: -0.005 min
Response: 1497302
Conc: 14.78 ng/ml




#40 AR-1262-5

R.T.: 9.480 min
Delta R.T.: 0.001 min
Response: 1510740
Conc: 17.64 ng/ml



#40 AR-1262-5


R.T.: 8.086 min
Delta R.T.: -0.003 min
Response: 383711
Conc: 8.82 ng/ml



#41 AR-1268-1


R.T.: 8.748 min
Delta R.T.: 0.037 min
Response: 26051543
Conc: 77.03 ng/ml

Instrument :
ECD_O
ClientSampleId :




#41 AR-1268-1

R.T.: 7.525 min
Delta R.T.: -0.003 min
Response: 434742
Conc: 2.57 ng/ml




#42 AR-1268-2

R.T.: 8.865 min
Delta R.T.: 0.056 min
Response: 41216
Conc: 0.14 ng/ml



#42 AR-1268-2


R.T.: 7.587 min
Delta R.T.: -0.006 min
Response: 1497302
Conc: 9.77 ng/ml



#43 AR-1268-3


R.T.: 9.040 min
Delta R.T.: -0.002 min
Response: 72539
Conc: 0.27 ng/ml

Instrument :
ECD_O
ClientSampleId :




#43 AR-1268-3

R.T.: 0.000 min
Exp R.T. : 7.799 min
Response: 0
Conc: N.D.




#44 AR-1268-4

R.T.: 9.480 min
Delta R.T.: 0.000 min
Response: 1510740
Conc: 16.09 ng/ml



#44 AR-1268-4


R.T.: 8.086 min
Delta R.T.: -0.003 min
Response: 383711
Conc: 7.68 ng/ml



#45 AR-1268-5

R.T.: 9.908 min
Delta R.T.: 0.002 min
Response: 686255
Conc: 0.82 ng/ml

Instrument :
ECD_O
ClientSampleId :



#45 AR-1268-5

R.T.: 8.378 min
Delta R.T.: -0.003 min
Response: 291645
Conc: 0.74 ng/ml