

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
 Data File : P0094751.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 19:34
 Operator : YP/AJ
 Sample : 02588-11
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ZONE-D-3-6-05022023

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/05/2023
 Supervised By :Ankita Jodhani 05/05/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 21:13:49 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.415	3.616	71552576	28013904	20.772	20.936
2)	SA Decachlor...	10.252	8.628	44973821	19277778	18.532	17.127
Target Compounds							
26)	L6 AR-1254-1	6.458	5.498	6527733	3785482	51.522	52.220m
27)	L6 AR-1254-2	6.679	5.644	9320824	2463743	48.766	37.759m
28)	L6 AR-1254-3	7.049	6.049	9065534	3562634	48.063	35.878m#
29)	L6 AR-1254-4	7.337	6.279	3414762	1606186	30.150	32.541
30)	L6 AR-1254-5	7.760	6.697	10865042	6820419	77.581m	81.069
31)	L7 AR-1260-1	7.214	6.180	6965232	4145099	46.977	58.879 #
32)	L7 AR-1260-2	7.472	6.370	12601462	3900397	77.708	48.789 #
33)	L7 AR-1260-3	7.838	6.522	4717761	2990107	36.087	39.761
34)	L7 AR-1260-4	8.064	6.996	6720411	1917345	48.060	30.629 #
35)	L7 AR-1260-5	8.393	7.239	8477000	5466462	34.649	43.371 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050423\
Data File : PO094751.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 19:34
Operator : YP/AJ
Sample : 02588-11
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ZONE-D-3-6-05022023

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/05/2023
Supervised By :Ankita Jodhani 05/05/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 21:13:49 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

