

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
 Data File : P0094722.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 11:06
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 18:20:16 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.419	3.618	162.3E6	64271682	47.129	48.033
2) SA Decachlor...	10.260	8.633	116.5E6	52262146	48.007	46.431
Target Compounds						
16) L4 AR-1242-1	5.597	4.699	37765013	15698972	452.076	474.031
17) L4 AR-1242-2	5.620	4.716	55979751	23892090	445.283	484.682
18) L4 AR-1242-3	5.682	4.894	36717082	12718051	450.880	488.802
19) L4 AR-1242-4	5.781	4.977	28684985	14207390	467.070	479.962
20) L4 AR-1242-5	6.526	5.500	29536761	15348506	458.978	485.789



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050423\
Data File : PO094722.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 11:06
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 18:20:16 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.419 min
Delta R.T.: 0.003 min
Response: 162342006
Conc: 47.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.618 min
Delta R.T.: 0.000 min
Response: 64271682
Conc: 48.03 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.260 min
Delta R.T.: 0.008 min
Response: 116504902
Conc: 48.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.633 min
Delta R.T.: 0.001 min
Response: 52262146
Conc: 46.43 ng/ml



#16 AR-1242-1


R.T.: 5.597 min
Delta R.T.: 0.003 min
Response: 37765013
Conc: 452.08 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.699 min
Delta R.T.: -0.001 min
Response: 15698972
Conc: 474.03 ng/ml




#17 AR-1242-2

R.T.: 5.620 min
Delta R.T.: 0.003 min
Response: 55979751
Conc: 445.28 ng/ml



#17 AR-1242-2


R.T.: 4.716 min
Delta R.T.: -0.002 min
Response: 23892090
Conc: 484.68 ng/ml



#18 AR-1242-3


R.T.: 5.682 min
Delta R.T.: 0.003 min
Response: 36717082
Conc: 450.88 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.894 min
Delta R.T.: 0.000 min
Response: 12718051
Conc: 488.80 ng/ml




#19 AR-1242-4

R.T.: 5.781 min
Delta R.T.: 0.003 min
Response: 28684985
Conc: 467.07 ng/ml



#19 AR-1242-4


R.T.: 4.977 min
Delta R.T.: -0.001 min
Response: 14207390
Conc: 479.96 ng/ml



#20 AR-1242-5

R.T.: 6.526 min
Delta R.T.: 0.003 min
Response: 29536761
Conc: 458.98 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.500 min
Delta R.T.: -0.001 min
Response: 15348506
Conc: 485.79 ng/ml