

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
 Data File : P0094736.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 04 May 2023 15:18
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 04 18:24:30 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.414	3.616	165.0E6	64876408	47.901	48.485
2)	SA Decachlor...	10.254	8.628	120.7E6	53800686	49.738	47.798
Target Compounds							
3)	L1 AR-1016-1	5.593	4.697	47673367	19445058	467.992	490.273
4)	L1 AR-1016-2	5.616	4.714	70801654	28783652	467.936	486.899
5)	L1 AR-1016-3	5.678	4.890	45648892	15655690	460.429	498.721
6)	L1 AR-1016-4	5.777	4.933	34122102	14281414	458.437	493.887
7)	L1 AR-1016-5	6.075	5.146	44796063	18184983	536.913	496.879
31)	L7 AR-1260-1	7.215	6.180	67466578	33314170	455.027	473.211
32)	L7 AR-1260-2	7.475	6.369	75536922	38293504	465.805	479.007
33)	L7 AR-1260-3	7.838	6.522	60716448	35663148	464.430	474.226
34)	L7 AR-1260-4	8.067	6.996	66370934	29977139	474.645	478.873
35)	L7 AR-1260-5	8.395	7.240	118.9E6	62446192	485.803	495.453



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050423\
Data File : P0094736.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 04 May 2023 15:18
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 04 18:24:30 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.002 min
Response: 165001866
Conc: 47.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.002 min
Response: 64876408
Conc: 48.49 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.254 min
Delta R.T.: 0.002 min
Response: 120707746
Conc: 49.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: -0.004 min
Response: 53800686
Conc: 47.80 ng/ml



#3 AR-1016-1


R.T.: 5.593 min
Delta R.T.: 0.000 min
Response: 47673367
Conc: 467.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.697 min
Delta R.T.: -0.003 min
Response: 19445058
Conc: 490.27 ng/ml




#4 AR-1016-2

R.T.: 5.616 min
Delta R.T.: 0.000 min
Response: 70801654
Conc: 467.94 ng/ml



#4 AR-1016-2


R.T.: 4.714 min
Delta R.T.: -0.003 min
Response: 28783652
Conc: 486.90 ng/ml



#5 AR-1016-3


R.T.: 5.678 min
Delta R.T.: -0.001 min
Response: 45648892
Conc: 460.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.890 min
Delta R.T.: -0.003 min
Response: 15655690
Conc: 498.72 ng/ml




#6 AR-1016-4

R.T.: 5.777 min
Delta R.T.: 0.000 min
Response: 34122102
Conc: 458.44 ng/ml



#6 AR-1016-4


R.T.: 4.933 min
Delta R.T.: -0.003 min
Response: 14281414
Conc: 493.89 ng/ml



#7 AR-1016-5


R.T.: 6.075 min
Delta R.T.: 0.000 min
Response: 44796063
Conc: 536.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.146 min
Delta R.T.: -0.003 min
Response: 18184983
Conc: 496.88 ng/ml




#31 AR-1260-1

R.T.: 7.215 min
Delta R.T.: 0.001 min
Response: 67466578
Conc: 455.03 ng/ml



#31 AR-1260-1


R.T.: 6.180 min
Delta R.T.: -0.004 min
Response: 33314170
Conc: 473.21 ng/ml



#32 AR-1260-2


R.T.: 7.475 min
Delta R.T.: 0.000 min
Response: 75536922
Conc: 465.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.369 min
Delta R.T.: -0.003 min
Response: 38293504
Conc: 479.01 ng/ml




#33 AR-1260-3

R.T.: 7.838 min
Delta R.T.: 0.000 min
Response: 60716448
Conc: 464.43 ng/ml



#33 AR-1260-3


R.T.: 6.522 min
Delta R.T.: -0.003 min
Response: 35663148
Conc: 474.23 ng/ml



#34 AR-1260-4


R.T.: 8.067 min
Delta R.T.: 0.000 min
Response: 66370934
Conc: 474.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 6.996 min
Delta R.T.: -0.004 min
Response: 29977139
Conc: 478.87 ng/ml



#35 AR-1260-5

R.T.: 8.395 min
Delta R.T.: 0.000 min
Response: 118852158
Conc: 485.80 ng/ml



#35 AR-1260-5

R.T.: 7.240 min
Delta R.T.: -0.003 min
Response: 62446192
Conc: 495.45 ng/ml