

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050523\
 Data File : P0094761.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 05 May 2023 10:24
 Operator : YP/AJ
 Sample : 02588-03
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ZONE-B-6-9-05012023

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/08/2023
 Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 05 21:19:50 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.616	67314891	26575136	19.542	19.861
2) SA Decachlor...	10.250	8.627	48952273	22281330	20.171	19.795
Target Compounds						
26) L6 AR-1254-1	6.456	5.497	5824050	3178564	45.968	43.848m
27) L6 AR-1254-2	6.677	5.645	9550112	2724730	49.965	41.759m
28) L6 AR-1254-3	7.047	6.048	7923851	3642531	42.010	36.682m
29) L6 AR-1254-4	7.336	6.278	3681538	1965561	32.505	39.822m
30) L6 AR-1254-5	7.759	6.697	7959258	3967946	56.833	47.164

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050523\
Data File : PO094761.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 05 May 2023 10:24
Operator : YP/AJ
Sample : 02588-03
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ZONE-B-6-9-05012023

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/08/2023
Supervised By :Ankita Jodhani 05/08/2023

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 05 21:19:50 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

