

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050823\
 Data File : P0094846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 04:31
 Operator : YP/AJ
 Sample : AR1242CCC500
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 05:25:48 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.414	3.615	165.9E6	66456112	48.167	49.666
2) SA Decachlor...	10.254	8.628	115.3E6	54725773	47.505	48.620
Target Compounds						
16) L4 AR-1242-1	5.592	4.696	38750304	15718246	463.870	474.613
17) L4 AR-1242-2	5.614	4.714	58599844	24299055	466.125	492.938
18) L4 AR-1242-3	5.678	4.890	37556591	12919298	461.189	496.536
19) L4 AR-1242-4	5.776	4.974	28188017	14515754	458.978	490.380
20) L4 AR-1242-5	6.522	5.497	29948892	15416430	465.382	487.938



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050823\
Data File : PO094846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 04:31
Operator : YP/AJ
Sample : AR1242CCC500
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 05:25:48 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.414 min
Delta R.T.: -0.002 min
Response: 165917854
Conc: 48.17 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.615 min
Delta R.T.: -0.002 min
Response: 66456112
Conc: 49.67 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.254 min
Delta R.T.: 0.002 min
Response: 115287950
Conc: 47.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.628 min
Delta R.T.: -0.003 min
Response: 54725773
Conc: 48.62 ng/ml



#16 AR-1242-1


R.T.: 5.592 min
Delta R.T.: -0.002 min
Response: 38750304
Conc: 463.87 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#16 AR-1242-1

R.T.: 4.696 min
Delta R.T.: -0.004 min
Response: 15718246
Conc: 474.61 ng/ml




#17 AR-1242-2

R.T.: 5.614 min
Delta R.T.: -0.002 min
Response: 58599844
Conc: 466.12 ng/ml



#17 AR-1242-2


R.T.: 4.714 min
Delta R.T.: -0.004 min
Response: 24299055
Conc: 492.94 ng/ml



#18 AR-1242-3


R.T.: 5.678 min
Delta R.T.: -0.002 min
Response: 37556591
Conc: 461.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500




#18 AR-1242-3

R.T.: 4.890 min
Delta R.T.: -0.005 min
Response: 12919298
Conc: 496.54 ng/ml




#19 AR-1242-4

R.T.: 5.776 min
Delta R.T.: -0.002 min
Response: 28188017
Conc: 458.98 ng/ml



#19 AR-1242-4


R.T.: 4.974 min
Delta R.T.: -0.005 min
Response: 14515754
Conc: 490.38 ng/ml



#20 AR-1242-5

R.T.: 6.522 min
Delta R.T.: -0.001 min
Response: 29948892
Conc: 465.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242CCC500



#20 AR-1242-5

R.T.: 5.497 min
Delta R.T.: -0.005 min
Response: 15416430
Conc: 487.94 ng/ml