

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050923\
 Data File : P0094874.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2023 12:59
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 09 21:13:19 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.415	3.616	162.2E6	64725128	47.085	48.372
2) SA Decachlor...	10.256	8.627	117.5E6	55309362	48.423	49.139
Target Compounds						
26) L6 AR-1254-1	6.457	5.497	55294137	34391153	436.424	474.420
27) L6 AR-1254-2	6.678	5.646	85958338	30745200	449.727	471.197
28) L6 AR-1254-3	7.048	6.049	83838812	45525214	444.495	458.462
29) L6 AR-1254-4	7.338	6.279	48888908	22340418	431.654	452.610
30) L6 AR-1254-5	7.762	6.697	74432688	38779168	531.483	460.936



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050923\
Data File : PO094874.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2023 12:59
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 09 21:13:19 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: 0.000 min
Response: 162190040
Conc: 47.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.616 min
Delta R.T.: -0.002 min
Response: 64725128
Conc: 48.37 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.256 min
Delta R.T.: 0.004 min
Response: 117515487
Conc: 48.42 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.627 min
Delta R.T.: -0.004 min
Response: 55309362
Conc: 49.14 ng/ml



#26 AR-1254-1


R.T.: 6.457 min
Delta R.T.: -0.006 min
Response: 55294137
Conc: 436.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.497 min
Delta R.T.: -0.007 min
Response: 34391153
Conc: 474.42 ng/ml




#27 AR-1254-2

R.T.: 6.678 min
Delta R.T.: -0.007 min
Response: 85958338
Conc: 449.73 ng/ml



#27 AR-1254-2


R.T.: 5.646 min
Delta R.T.: -0.006 min
Response: 30745200
Conc: 471.20 ng/ml



#28 AR-1254-3


R.T.: 7.048 min
Delta R.T.: -0.008 min
Response: 83838812
Conc: 444.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.049 min
Delta R.T.: -0.007 min
Response: 45525214
Conc: 458.46 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: -0.007 min
Response: 48888908
Conc: 431.65 ng/ml



#29 AR-1254-4


R.T.: 6.279 min
Delta R.T.: -0.007 min
Response: 22340418
Conc: 452.61 ng/ml



#30 AR-1254-5

R.T.: 7.762 min
Delta R.T.: -0.008 min
Response: 74432688
Conc: 531.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.697 min
Delta R.T.: -0.008 min
Response: 38779168
Conc: 460.94 ng/ml