

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0050924\
 Data File : P0103524.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 09 May 2024 14:09
 Operator : YP/AJ
 Sample : P2421-22 10X
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 CABLE-27A

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 05/10/2024
 Supervised By :Ankita Jodhani 05/10/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 01:06:10 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0050624.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 07 04:55:35 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.452	3.588	10999702	4602354	1.429m	1.827 #
2) SA Decachlor...	10.212	8.552	8034534	2036436	1.695	1.565

Target Compounds

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO050924\
Data File : PO103524.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 09 May 2024 14:09
Operator : YP/AJ
Sample : P2421-22 10X
Misc :
ALS Vial : 16 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampleId :
CABLE-27A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/10/2024
Supervised By :Ankita Jodhani 05/10/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 01:06:10 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO050624.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 07 04:55:35 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.452 min
Delta R.T.: 0.001 min
Response: 10999702
Conc: 1.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
CABLE-27A


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 05/10/2024
Supervised By :Ankita Jodhani 05/10/2024




#1 Tetrachloro-m-xylene

R.T.: 3.588 min
Delta R.T.: -0.004 min
Response: 4602354
Conc: 1.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.212 min
Delta R.T.: 0.010 min
Response: 8034534
Conc: 1.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.552 min
Delta R.T.: -0.004 min
Response: 2036436
Conc: 1.57 ng/ml