

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051023\
 Data File : P0094923.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 May 2023 16:23
 Operator : YP/AJ
 Sample : AR1254CCC500
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 10 21:31:57 2023
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0042623.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Mon May 01 11:12:50 2023
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.415	3.617	165.3E6	67255979	47.997	50.264
2)	SA Decachlor...	10.252	8.627	123.0E6	57526110	50.673	51.108
Target Compounds							
26)	L6 AR-1254-1	6.457	5.498	59820261	35289088	472.147	486.807
27)	L6 AR-1254-2	6.679	5.646	88517697	31607158	463.118	484.407
28)	L6 AR-1254-3	7.048	6.049	85575345	47576224	453.701	479.117
29)	L6 AR-1254-4	7.338	6.279	49365123	23532154	435.859	476.754
30)	L6 AR-1254-5	7.761	6.697	63309215	40806174	452.056	485.029



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051023\
Data File : PO094923.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 May 2023 16:23
Operator : YP/AJ
Sample : AR1254CCC500
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 10 21:31:57 2023
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO042623.M
Quant Title : GC EXTRACTABLES
QLast Update : Mon May 01 11:12:50 2023
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.415 min
Delta R.T.: 0.000 min
Response: 165330365
Conc: 48.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.617 min
Delta R.T.: -0.001 min
Response: 67255979
Conc: 50.26 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.252 min
Delta R.T.: 0.000 min
Response: 122976101
Conc: 50.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.627 min
Delta R.T.: -0.004 min
Response: 57526110
Conc: 51.11 ng/ml



#26 AR-1254-1


R.T.: 6.457 min
Delta R.T.: -0.007 min
Response: 59820261
Conc: 472.15 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#26 AR-1254-1

R.T.: 5.498 min
Delta R.T.: -0.006 min
Response: 35289088
Conc: 486.81 ng/ml




#27 AR-1254-2

R.T.: 6.679 min
Delta R.T.: -0.007 min
Response: 88517697
Conc: 463.12 ng/ml



#27 AR-1254-2


R.T.: 5.646 min
Delta R.T.: -0.006 min
Response: 31607158
Conc: 484.41 ng/ml



#28 AR-1254-3


R.T.: 7.048 min
Delta R.T.: -0.008 min
Response: 85575345
Conc: 453.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500




#28 AR-1254-3

R.T.: 6.049 min
Delta R.T.: -0.007 min
Response: 47576224
Conc: 479.12 ng/ml




#29 AR-1254-4

R.T.: 7.338 min
Delta R.T.: -0.007 min
Response: 49365123
Conc: 435.86 ng/ml



#29 AR-1254-4


R.T.: 6.279 min
Delta R.T.: -0.007 min
Response: 23532154
Conc: 476.75 ng/ml



#30 AR-1254-5

R.T.: 7.761 min
Delta R.T.: -0.009 min
Response: 63309215
Conc: 452.06 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254CCC500



#30 AR-1254-5

R.T.: 6.697 min
Delta R.T.: -0.008 min
Response: 40806174
Conc: 485.03 ng/ml