

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051121\
 Data File : P0077698.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 May 2021 19:50
 Operator : DD\AJ
 Sample : M2262-07
 Misc : AR1268 LOD 25 PPB
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 LOD-MDL-WATER-SOIL-01-QT2-202

Manual Integrations
 APPROVED

Ankita
 5/13/2021 10:47:06 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 23:02:06 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 15:14:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.366	3.645	1136078	291708	16.365	17.349
2) SA Decachlor...	9.991	8.848	995535	409464	18.947	23.218
Target Compounds						
41) L9 AR-1268-1	8.759	7.802	166897	65025	22.295m	26.497
42) L9 AR-1268-2	8.841	7.865	144304	59316	21.483	26.835
43) L9 AR-1268-3	9.026	8.070	112791	52660	20.185	28.015 #
44) L9 AR-1268-4	9.392	8.359	44482	22331	18.288	26.174 #
45) L9 AR-1268-5	9.726	8.630	403142	163341	21.514	27.039 #

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051121\
Data File : P0077698.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 May 2021 19:50
Operator : DD\AJ
Sample : M2262-07
Misc : AR1268 LOD 25 PPB
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOD-MDL-WATER-SOIL-01-QT2-202

Manual Integrations
APPROVED

Ankita
5/13/2021 10:47:06 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 23:02:06 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 15:14:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

