

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_0\Data\P0051121\
 Data File : P0077735.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 12 May 2021 7:52
 Operator : DD\AJ
 Sample : M2262-08
 Misc : AR1232 LOQ 50 PPB
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 ECD_0
 ClientSampleId :
 LOQ-WATER-02-QT2-2021

Manual Integrations
 APPROVED

Ankita
 5/13/2021 10:49:44 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 12 13:35:02 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_0\methods\P0051121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 15:14:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.365	3.645	1196618	388903	17.237	23.130 #
2) SA Decachlor...	9.988	8.847	959930	517041	18.270	29.317 #
Target Compounds						
11) L3 AR-1232-1	4.797	4.083	75597	26595	52.665m	70.442 #
12) L3 AR-1232-2	5.369	4.898	68280	25339	94.107	74.219
13) L3 AR-1232-3	5.685	5.084	71289	13419	51.568	73.382 #
14) L3 AR-1232-4	5.855	5.182	40852	12460	59.612	75.208 #
15) L3 AR-1232-5	5.955	5.362	25588	11389	55.348	62.877

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051121\
Data File : PO077735.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 12 May 2021 7:52
Operator : DD\AJ
Sample : M2262-08
Misc : AR1232 LOQ 50 PPB
ALS Vial : 50 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
LOQ-WATER-02-QT2-2021

Manual Integrations
APPROVED

Ankita
5/13/2021 10:49:44 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 12 13:35:02 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 15:14:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

