

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051121\
 Data File : P0077674.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 May 2021 12:57
 Operator : DD\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Manual Integrations
 APPROVED

Ankita
 5/13/2021 10:46:22 AM

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 15:09:18 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 15:09:03 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.367	3.645	3336354	815495	50.000	50.000
2) SA Decachlor...	9.993	8.848	2561816	871703	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	7.951	6.971	1917651	323061	500.000	500.000
37) L8 AR-1262-2	8.489	7.519	3441951	1106444	500.000	500.000
38) L8 AR-1262-3	8.762	7.802	2229685	447389	500.000	500.000
39) L8 AR-1262-4	8.840	7.865	1631278	833910	493.606m	500.000
40) L8 AR-1262-5	9.393	8.361	1160283	402644	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO051121\
Data File : PO077674.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 May 2021 12:57
Operator : DD\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 13 Sample Multiplier: 1



Instrument :
ECD_O
ClientSampled :
AR1262ICC500


Manual Integrations
APPROVED

Ankita
5/13/2021 10:46:22 AM

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 15:09:18 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO051121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 15:09:03 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.367 min
Delta R.T.: 0.000 min
Response: 3336354
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Ankita
5/13/2021 10:46:22 AM




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: 0.000 min
Response: 815495
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 9.993 min
Delta R.T.: 0.000 min
Response: 2561816
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.848 min
Delta R.T.: 0.000 min
Response: 871703
Conc: 50.00 ng/ml




#36 AR-1262-1

R.T.: 7.951 min
Delta R.T.: 0.000 min
Response: 1917651
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Ankita
5/13/2021 10:46:22 AM




#36 AR-1262-1

R.T.: 6.971 min
Delta R.T.: 0.000 min
Response: 323061
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.489 min
Delta R.T.: 0.000 min
Response: 3441951
Conc: 500.00 ng/ml



#37 AR-1262-2

R.T.: 7.519 min
Delta R.T.: 0.000 min
Response: 1106444
Conc: 500.00 ng/ml




#38 AR-1262-3

R.T.: 8.762 min
Delta R.T.: 0.000 min
Response: 2229685
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500


Manual Integrations
APPROVED

Ankita
5/13/2021 10:46:22 AM




#38 AR-1262-3

R.T.: 7.802 min
Delta R.T.: 0.000 min
Response: 447389
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 8.840 min
Delta R.T.: 0.000 min
Response: 1631278
Conc: 493.61 ng/ml m



#39 AR-1262-4

R.T.: 7.865 min
Delta R.T.: 0.000 min
Response: 833910
Conc: 500.00 ng/ml




#40 AR-1262-5

R.T.: 9.393 min
Delta R.T.: 0.000 min
Response: 1160283
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Manual Integrations
APPROVED

Ankita
5/13/2021 10:46:22 AM



#40 AR-1262-5

R.T.: 8.361 min
Delta R.T.: 0.000 min
Response: 402644
Conc: 500.00 ng/ml