

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051121\
 Data File : P0077700.D
 Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
 Acq On : 11 May 2021 20:24
 Operator : DD\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 11 23:03:01 2021
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051121.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue May 11 15:14:32 2021
 Response via : Initial Calibration
 Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.367	3.645	1348328	352788	19.423	20.982
2) SA Decachlor...	9.991	8.848	1080945	447574	20.573	25.378

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0051121\
Data File : P0077700.D
Signal(s) : Signal #1: ECD1A.CH Signal #2: ECD2B.CH
Acq On : 11 May 2021 20:24
Operator : DD\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 11 23:03:01 2021
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0051121.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue May 11 15:14:32 2021
Response via : Initial Calibration
Integrator: ChemStation 6890 Scale Mode: Small noise peaks clipped

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.367 min
Delta R.T.: -0.002 min
Response: 1348328
Conc: 19.42 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.645 min
Delta R.T.: -0.002 min
Response: 352788
Conc: 20.98 ng/ml



#2 Decachlorobiphenyl

R.T.: 9.991 min
Delta R.T.: -0.001 min
Response: 1080945
Conc: 20.57 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.848 min
Delta R.T.: -0.002 min
Response: 447574
Conc: 25.38 ng/ml